

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW072720\
 Data File : VW015971.D
 Acq On : 27 Jul 2020 16:59
 Operator : SY/VA
 Sample : L3458-07
 Misc : 4.89G/10ML/MSVOA W/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 GAY35

Quant Time: Jul 28 02:52:41 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM072020S.M
 Quant Title : VOC Analysis
 QLast Update : Tue Jul 28 01:36:11 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.85	114	442250	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.63	117	414095	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.56	152	195273	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.36	65	102981	16.50	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	66.00%
7) Chloroethane-d5	2.89	69	83301	16.44	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	65.76%
10) 1,1-Dichloroethene-d2	4.01	63	155257	12.31	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	49.24%
20) 2-Butanone-d5	7.09	46	111785	68.19	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	136.38%#
24) Chloroform-d	7.65	84	214091	17.24	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	68.96%
26) 1,2-Dichloroethane-d4	8.31	65	136623	20.04	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	80.16%
29) Benzene-d6	8.28	84	430157	17.07	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	68.28%
33) 1,2-Dichloropropane-d6	9.28	67	138992	18.39	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	73.56%
37) Toluene-d8	10.33	98	378682	16.59	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	66.36%
38) trans-1,3-Dichloropropene-	10.58	79	62164	19.30	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	77.20%
39) 2-Hexanone-d5	10.93	63	90001	71.53	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	143.06%#
48) 1,1,1,2,2-Tetrachloroethane-	12.69	84	157477	27.11	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	108.44%
61) 1,2-Dichlorobenzene-d4	13.86	152	142715	18.74	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	74.96%#

Target Compounds

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
13) Acetone	4.14	43	41529	30.167	ug/L	96
15) Methyl Acetate	4.68	43	6444	2.483	ug/L	98
16) Methylene chloride	4.92	84	17190	2.372	ug/L	92
21) 2-Butanone	7.18	43	10812	6.351	ug/L	98
22) cis-1,2-Dichloroethene	7.17	96	2665	0.424	ug/L	84
35) Trichloroethene	9.10	95	14221	2.137	ug/L	98
44) Toluene	10.40	91	10174	0.386	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW072720\
Data File : VW015971.D
Acq On : 27 Jul 2020 16:59
Operator : SY/VA
Sample : L3458-07
Misc : 4.89G/10ML/MSVOA W/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
GAY35

Quant Time: Jul 28 02:52:41 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM072020S.M
Quant Title : VOC Analysis
QLast Update : Tue Jul 28 01:36:11 2020
Response via : Initial Calibration

