


Methylene Chloride

Response Ratio




Resp Ratio = $5.55 \times 10^{-1} \times \text{Amt} + 1.20 \times 10^{-1}$
Coef of Det (r^2) = 0.999 Curve Fit: Linear

Method Name: Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W080618S.M
Calibration Table Last Updated: Tue Aug 07 16:59:03 2018

2-Chloroethyl Vinyl ether

Response Ratio




Resp Ratio = $1.90 \times 10^{-1} \times \text{Amt} - 3.41 \times 10^{-2}$
Coef of Det (r^2) = 0.999 Curve Fit: Linear

Method Name: Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W080618S.M
Calibration Table Last Updated: Tue Aug 07 16:59:03 2018

Naphthalene

Response Ratio



Resp Ratio = 2.37e+000 * Amt - 1.38e-001
Coef of Det (r^2) = 1.000 Curve Fit: Linear

Method Name: Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W080618S.M
Calibration Table Last Updated: Tue Aug 07 16:59:03 2018