

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW081019\
 Data File : VW011786.D
 Acq On : 08 Aug 2019 23:47
 Operator : SY/VA
 Sample : K4228-04
 Misc : 5.60G/5ML/MSVOA W/SOIL
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 2S-F2-(0)

Quant Time: Aug 09 05:57:42 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W072719S.M
 Quant Title : SW846 8260
 QLast Update : Sat Jul 27 00:23:30 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	226675	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	374982	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	307547	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	121123	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	145020	55.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.44%	
35) Dibromofluoromethane	7.88	113	114998	54.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.32%	
50) Toluene-d8	10.32	98	457357	53.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.20%	
62) 4-Bromofluorobenzene	12.62	95	143525	46.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.48%	
Target Compounds						
16) Acetone	4.13	43	60363	72.905	ug/l	94
20) Methylene Chloride	4.92	84	16225	5.682	ug/l	96
68) m/p-Xylenes	11.84	106	5156	1.204	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW081019\
Data File : VW011786.D
Acq On : 08 Aug 2019 23:47
Operator : SY/VA
Sample : K4228-04
Misc : 5.60G/5ML/MSVOA W/SOIL
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
2S-F2-(0)

Quant Time: Aug 09 05:57:42 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W072719S.M
Quant Title : SW846 8260
QLast Update : Sat Jul 27 00:23:30 2019
Response via : Initial Calibration

