

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW081020\
 Data File : VW016113.D
 Acq On : 10 Aug 2020 20:41
 Operator : SY/VA
 Sample : L3598-09
 Misc : 6.80G/10ML/MSVOA W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :

Quant Time: Aug 11 02:24:35 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM081020S.M
 Quant Title : VOC Analysis
 QLast Update : Tue Aug 11 01:19:31 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.85	114	426578	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.63	117	383002	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.56	152	147096	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.35	65	124094	23.46	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	93.84%
7) Chloroethane-d5	2.89	69	91212	22.72	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	90.88%
10) 1,1-Dichloroethene-d2	4.01	63	193776	17.25	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	69.00%
20) 2-Butanone-d5	7.09	46	99792	71.71	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	143.42%#
24) Chloroform-d	7.65	84	272736	23.82	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	95.28%
26) 1,2-Dichloroethane-d4	8.30	65	171529	28.21	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	112.84%
29) Benzene-d6	8.27	84	526888	25.29	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	101.16%
33) 1,2-Dichloropropane-d6	9.27	67	160042	26.12	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	104.48%
37) Toluene-d8	10.32	98	475768	24.90	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	99.60%
38) trans-1,3-Dichloropropene-	10.58	79	74972	27.23	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	108.92%
39) 2-Hexanone-d5	10.93	63	84210	79.67	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	159.34%#
48) 1,1,2,2-Tetrachloroethane-	12.69	84	165906	33.26	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	133.04%#
61) 1,2-Dichlorobenzene-d4	13.85	152	142183	27.27	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	109.08%

Target Compounds

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
13) Acetone	4.14	43	52804	50.157	ug/L	97
16) Methylene chloride	4.90	84	11982m	1.779	ug/L	
21) 2-Butanone	7.19	43	6705	4.019	ug/L	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW081020\
Data File : VW016113.D
Acq On : 10 Aug 2020 20:41
Operator : SY/VA
Sample : L3598-09
Misc : 6.80G/10ML/MSVOA W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :

Quant Time: Aug 11 02:24:35 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM081020S.M
Quant Title : VOC Analysis
QLast Update : Tue Aug 11 01:19:31 2020
Response via : Initial Calibration

