

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW081020\
 Data File : VW016109.D
 Acq On : 10 Aug 2020 19:05
 Operator : SY/VA
 Sample : L3598-12
 Misc : 6.62G/10ML/MSVOA W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 GAY68

Quant Time: Aug 11 02:08:14 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM081020S.M
 Quant Title : VOC Analysis
 QLast Update : Tue Aug 11 01:19:31 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.85	114	412694	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.63	117	391200	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.56	152	158986	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.36	65	114484	22.37	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	89.48%
7) Chloroethane-d5	2.89	69	83256	21.44	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	85.76%
10) 1,1-Dichloroethene-d2	4.01	63	178208	16.40	ug/L	-0.01
Spiked Amount	25.000	Range	45 - 110	Recovery	=	65.60%
20) 2-Butanone-d5	7.09	46	82192	61.05	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	122.10%
24) Chloroform-d	7.65	84	245525	22.17	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	88.68%
26) 1,2-Dichloroethane-d4	8.31	65	152605	25.94	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	103.76%
29) Benzene-d6	8.27	84	477056	22.42	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	89.68%
33) 1,2-Dichloropropane-d6	9.27	67	144559	23.10	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	92.40%
37) Toluene-d8	10.32	98	437403	22.41	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	89.64%
38) trans-1,3-Dichloropropene-	10.58	79	68873	24.49	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	97.96%
39) 2-Hexanone-d5	10.93	63	66653	61.74	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	123.48%
48) 1,1,2,2-Tetrachloroethane-	12.69	84	142888	28.05	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	112.20%
61) 1,2-Dichlorobenzene-d4	13.85	152	145113	25.75	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	103.00%

Target Compounds					Ovalue
13) Acetone	4.14	43	7466	7.330 ug/L	99
16) Methylene chloride	4.92	84	10520	1.614 ug/L	95
34) Benzene	8.32	78	10346	0.459 ug/L	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW081020\
Data File : VW016109.D
Acq On : 10 Aug 2020 19:05
Operator : SY/VA
Sample : L3598-12
Misc : 6.62G/10ML/MSVOA W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
GAY68

Quant Time: Aug 11 02:08:14 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM081020S.M
Quant Title : VOC Analysis
QLast Update : Tue Aug 11 01:19:31 2020
Response via : Initial Calibration

