

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW081820\
 Data File : VW016242.D
 Acq On : 18 Aug 2020 15:46
 Operator : SY/VA
 Sample : L3709-11
 Misc : 6.71G/10ML/MSVOA W/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BFWQ2

Quant Time: Aug 19 06:57:37 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM081020S.M
 Quant Title : VOC Analysis
 QLast Update : Wed Aug 19 05:49:31 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.85	114	248949	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.63	117	219701	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.56	152	72929	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.35	65	58977	19.11	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	76.44%
7) Chloroethane-d5	2.89	69	48358	20.64	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	82.56%
10) 1,1-Dichloroethene-d2	4.01	63	99510	15.18	ug/L	-0.01
Spiked Amount	25.000	Range	45 - 110	Recovery	=	60.72%
20) 2-Butanone-d5	7.09	46	30979	38.14	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	76.28%
24) Chloroform-d	7.65	84	146273	21.89	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	87.56%
26) 1,2-Dichloroethane-d4	8.30	65	71920	20.27	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	81.08%
29) Benzene-d6	8.27	84	275459	23.05	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	92.20%
33) 1,2-Dichloropropane-d6	9.27	67	83790	23.84	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	95.36%
37) Toluene-d8	10.32	98	241378	22.02	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	88.08%
38) trans-1,3-Dichloropropene-	10.58	79	31928	20.21	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	80.84%
39) 2-Hexanone-d5	10.93	63	25858	42.65	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	85.30%
48) 1,1,2,2-Tetrachloroethane-	12.69	84	64093	22.40	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	89.60%
61) 1,2-Dichlorobenzene-d4	13.86	152	62057	24.01	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	96.04%

Target Compounds					Ovalue
13) Acetone	4.14	43	4716	7.676 ug/L	90
16) Methylene chloride	4.91	84	14250	3.625 ug/L	96
47) Tetrachloroethene	10.87	164	85345	33.417 ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW081820\
Data File : VW016242.D
Acq On : 18 Aug 2020 15:46
Operator : SY/VA
Sample : L3709-11
Misc : 6.71G/10ML/MSVOA W/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BFWQ2

Quant Time: Aug 19 06:57:37 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM081020S.M
Quant Title : VOC Analysis
QLast Update : Wed Aug 19 05:49:31 2020
Response via : Initial Calibration

