

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW082120\
 Data File : VW016282.D
 Acq On : 21 Aug 2020 00:45
 Operator : SY/VA
 Sample : L3712-14RE
 Misc : 6.84G/5ML/MSVOA W/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 CHRT-26982RE

Quant Time: Aug 21 03:51:27 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W081420S.M
 Quant Title : SW846 8260
 QLast Update : Mon Aug 17 08:01:41 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	182324	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	357480	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	257687	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	63801	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	8.31	65	108558	61.06	ug/l	0.00
Spiked Amount	50.000		Recovery	=	122.12%	
35) Dibromofluoromethane	7.88	113	112219	50.88	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.76%	
50) Toluene-d8	10.32	98	385520	45.15	ug/l	0.00
Spiked Amount	50.000		Recovery	=	90.30%	
62) 4-Bromofluorobenzene	12.62	95	98557	32.29	ug/l	0.00
Spiked Amount	50.000		Recovery	=	64.58%	

Target Compounds


					Qvalue	
3) Chloromethane	2.32	50	1506	1.141	ug/l	92
11) Tert butyl alcohol	5.15	59	12018	127.002	ug/l #	81
16) Acetone	4.12	43	5960	17.498	ug/l	100
20) Methylene Chloride	4.92	84	27927	9.163	ug/l	93
25) 2-Butanone	7.18	43	1659	3.634	ug/l #	72
31) Cyclohexane	7.95	56	4885	1.505	ug/l #	1
36) 1,1-Dichloropropene	7.95	75	14291	4.058	ug/l #	52
38) Carbon Tetrachloride	7.95	117	19553	5.564	ug/l #	15
52) Toluene	10.39	92	11900	1.876	ug/l	99
55) 1,1,2-Trichloroethane	10.76	97	2381	1.353	ug/l #	26
59) 2-Hexanone	10.94	43	3519	4.477	ug/l #	29
68) m/p-Xylenes	11.83	106	13815	3.786	ug/l	99
69) o-Xylene	12.16	106	6028	1.854	ug/l	85
74) N-amyl acetate	12.26	43	10112	11.614	ug/l #	67
75) 1,1,2,2-Tetrachloroethane	12.69	83	4357	5.356	ug/l #	73
76) 1,2,3-Trichloropropane	12.62	75	46011	78.415	ug/l #	1
80) 1,3,5-Trimethylbenzene	12.94	105	11246	2.695	ug/l	93
81) trans-1,4-Dichloro-2-buten	12.62	75	46011	188.090	ug/l #	14
83) tert-Butylbenzene	13.44	119	6596	1.928	ug/l	85
84) 1,2,4-Trimethylbenzene	13.26	105	21299	5.157	ug/l	97
86) p-Isopropyltoluene	13.50	119	27462	6.108	ug/l	96
89) n-Butylbenzene	13.79	91	9576	2.302	ug/l #	1
90) Hexachloroethane	14.10	117	2833	3.376	ug/l #	39
92) 1,2-Dibromo-3-Chloropropan	14.50	75	363	2.903	ug/l #	19
93) 1,2,4-Trichlorobenzene	15.13	180	1481	1.323	ug/l #	1
95) Naphthalene	15.36	128	5987	5.945	ug/l #	82
96) 1,2,3-Trichlorobenzene	15.58	180	8461	8.799	ug/l #	12


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW082120\
Data File : VW016282.D
Acq On : 21 Aug 2020 00:45
Operator : SY/VA
Sample : L3712-14RE
Misc : 6.84G/5ML/MSVOA W/SOIL
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
CHRT-26982RE

Quant Time: Aug 21 03:51:27 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W081420S.M
Quant Title : SW846 8260
QLast Update : Mon Aug 17 08:01:41 2020
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.95 min Scan# 991

Delta R.T. -0.01 min

Lab File: VW016282.D

Acq: 21 Aug 2020 00:45

Instrument :

MSVOA_W

ClientSampleId :


CHRT-26982RE

Tot Ion:168 Resp: 182324

Ion Ratio Lower Upper




168 100

99 63.2 48.3 72.5



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VW



#3

Chloromethane

Concen: 1.141 ug/l

RT: 2.32 min Scan# 68

Delta R.T. 0.10 min

Lab File: VW016282.D


Acq: 21 Aug 2020 00:45

Tgt Ion: 50 Resp: 1506

Ion Ratio Lower Upper



50 100


52 27.2 25.3 37.9



Abundance Ion 49.90 (49.60 to 50.60): VW

Ion 51.90 (51.60 to 52.60): VW









#11

Tert butyl alcohol
 Concen: 127.002 ug/l
 RT: 5.15 min Scan# 533
 Delta R.T. -0.07 min
 Lab File: VW016282.D
 Acq: 21 Aug 2020 00:45

Instrument :
 MSVOA_W
 ClientSampleId :
 CHRT-26982RE




Tot Ion: 59 Resp: 12018
 Ion Ratio Lower Upper
 59 100
 57 4.2 9.0 13.6#




#16

Acetone
 Concen: 17.498 ug/l
 RT: 4.12 min Scan# 364
 Delta R.T. -0.02 min
 Lab File: VW016282.D
 Acq: 21 Aug 2020 00:45

Tgt Ion: 43 Resp: 5960
 Ion Ratio Lower Upper
 43 100
 58 32.2 25.8 38.6









#20
Methvlene Chloride
Concen: 9.163 ug/l
RT: 4.92 min Scan# 494
Delta R.T. 0.00 min
Lab File: VW016282.D
Acq: 21 Aug 2020 00:45

Instrument :
MSVOA_W
ClientSampleId :
CHRT-26982RE

Tot Ion:	84	Resp:	27927
Ion	Ratio	Lower	Upper
84	100		
49	106.4	93.4	140.2
51	37.2	26.6	40.0
86	67.1	52.2	78.2




Abundance Ion 83.90 (83.60 to 84.60): VW
Ion 48.90 (48.60 to 49.60): VW
Ion 50.90 (50.60 to 51.60): VW
Ion 85.90 (85.60 to 86.60): VW






#25
2-Butanone
Concen: 3.634 ug/l
RT: 7.18 min Scan# 865
Delta R.T. -0.01 min
Lab File: VW016282.D
Acq: 21 Aug 2020 00:45

Tgt Ion:	43	Resp:	1659
Ion	Ratio	Lower	Upper
43	100		
72	14.8	24.2	36.2#



Abundance Ion 43.00 (42.70 to 43.70): VW
Ion 72.00 (71.70 to 72.70): VW





#31

Cyclohexane

Concen: 1.505 ug/l

RT: 7.95 min Scan# 991

Delta R.T. -0.01 min

Lab File: VW016282.D

Acq: 21 Aug 2020 00:45

Instrument :

MSVOA_W

ClientSampleId :

CHRT-26982RE


Tot Ion: 56 Resp: 4885

Ion Ratio Lower Upper




56 100

69 183.2 27.3 40.9#

84 136.9 73.5 110.3#



Abundance Ion 56.00 (55.70 to 56.70): VW
Ion 69.00 (68.70 to 69.70): VW
Ion 84.00 (83.70 to 84.70): VW



#33

1,2-Dichloroethane-d4

Concen: 61.056 ug/l

RT: 8.31 min Scan# 1050

Delta R.T. -0.01 min

Lab File: VW016282.D


Acq: 21 Aug 2020 00:45

Tgt Ion: 65 Resp: 108558



Ion Ratio Lower Upper


65 100

67 51.2 0.0 105.6



Abundance Ion 64.90 (64.60 to 65.60): VW
Ion 66.90 (66.60 to 67.60): VW





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.84 min Scan# 1138

Delta R.T. -0.01 min

Lab File: VW016282.D

Acq: 21 Aug 2020 00:45

Instrument :

MSVOA_W

ClientSampleId :

CHRT-26982RE


Tot Ion:114 Resp: 357480

Ion Ratio Lower Upper

114 100

63 19.6 0.0 38.4




88 16.6 0.0 32.8



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VW

Ion 87.90 (87.60 to 88.60): VW



#35

Dibromofluoromethane

Concen: 50.877 ug/l

RT: 7.88 min Scan# 981

Delta R.T. -0.01 min

Lab File: VW016282.D

Acq: 21 Aug 2020 00:45


Tgt Ion:113 Resp: 112219

Ion Ratio Lower Upper

113 100

111 101.1 83.5 125.3



192 18.6 14.3 21.5




Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V









#36
 1,1-Dichloropropene
 Concen: 4.058 ug/l
 RT: 7.95 min Scan# 991
 Delta R.T. -0.14 min
 Lab File: VW016282.D
 Acq: 21 Aug 2020 00:45




Instrument :
 MSVOA_W
 ClientSampleId :
 CHRT-26982RE


Tot Ion: 75 Resp: 14291
 Ion Ratio Lower Upper
 75 100
 110 11.4 18.1 54.3#
 77 0.2 24.7 37.1#



#38
 Carbon Tetrachloride
 Concen: 5.564 ug/l
 RT: 7.95 min Scan# 991
 Delta R.T. -0.13 min
 Lab File: VW016282.D
 Acq: 21 Aug 2020 00:45

Tgt Ion: 117 Resp: 19553
 Ion Ratio Lower Upper
 117 100
 119 3.9 76.6 115.0#
 121 0.8 25.4 38.0#









#50
Toluene-d8
Concen: 45.154 ug/l
RT: 10.32 min Scan# 1381
Delta R.T. -0.00 min
Lab File: VW016282.D
Acq: 21 Aug 2020 00:45




Instrument :
MSVOA_W
ClientSampleId :
CHRT-26982RE


Tot Ion: 98 Resp: 385520
Ion Ratio Lower Upper
98 100
100 66.9 53.5 80.3



#52
Toluene
Concen: 1.876 ug/l
RT: 10.39 min Scan# 1392
Delta R.T. -0.00 min
Lab File: VW016282.D
Acq: 21 Aug 2020 00:45

Tgt Ion: 92 Resp: 11900
Ion Ratio Lower Upper
92 100
91 167.6 133.3 199.9









#55
 1,1,2-Trichloroethane
 Concen: 1.353 ug/l
 RT: 10.76 min Scan# 1452
 Delta R.T. -0.03 min
 Lab File: VW016282.D
 Acq: 21 Aug 2020 00:45




Instrument :
 MSVOA_W
 ClientSampleId :
 CHRT-26982RE


Tot Ion: 97 Resp: 2381
 Ion Ratio Lower Upper
 97 100
 83 5.2 66.6 100.0#
 85 19.4 43.0 64.4#
 99 0.2 49.4 74.2#



#59
 2-Hexanone
 Concen: 4.477 ug/l
 RT: 10.94 min Scan# 1482
 Delta R.T. -0.03 min
 Lab File: VW016282.D
 Acq: 21 Aug 2020 00:45

Tgt Ion: 43 Resp: 3519
 Ion Ratio Lower Upper
 43 100
 58 5.0 28.8 86.5#









#62
4-Bromofluorobenzene
Concen: 32.294 ug/l
RT: 12.62 min Scan# 1758
Delta R.T. -0.00 min
Lab File: VW016282.D
Acq: 21 Aug 2020 00:45




Instrument :
MSVOA_W
ClientSampleId :
CHRT-26982RE


Tot Ion: 95 Resp: 98557
Ion Ratio Lower Upper
95 100
174 71.4 0.0 145.6
176 70.4 0.0 137.8



#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.63 min Scan# 1596
Delta R.T. -0.00 min
Lab File: VW016282.D
Acq: 21 Aug 2020 00:45

Tgt Ion: 117 Resp: 257687
Ion Ratio Lower Upper
117 100
82 55.7 43.6 65.4
119 33.1 25.4 38.0









#68
m/p-Xylenes
Concen: 3.786 ug/l
RT: 11.83 min Scan# 1629
Delta R.T. -0.01 min
Lab File: VW016282.D
Acq: 21 Aug 2020 00:45




Instrument :
MSVOA_W
ClientSampleId :
CHRT-26982RE


Tot Ion:106 Resp: 13815
Ion Ratio Lower Upper
106 100
91 200.2 161.4 242.2



#69
o-Xylene
Concen: 1.854 ug/l
RT: 12.16 min Scan# 1683
Delta R.T. -0.01 min
Lab File: VW016282.D
Acq: 21 Aug 2020 00:45

Tgt Ion:106 Resp: 6028
Ion Ratio Lower Upper
106 100
91 235.0 106.0 317.9





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.56 min Scan# 1912

Delta R.T. -0.00 min

Lab File: VW016282.D

Acq: 21 Aug 2020 00:45

Instrument :

MSVOA_W

ClientSampleId :

CHRT-26982RE


Tot Ion:152 Resp: 63801

Ion Ratio Lower Upper

152 100

115 54.7 30.9 92.5




150 137.8 0.0 337.8



Abundance Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V



#74

N-ethyl acetate

Concen: 11.614 ug/l

RT: 12.26 min Scan# 1698

Delta R.T. -0.02 min

Lab File: VW016282.D

Acq: 21 Aug 2020 00:45

Tgt Ion: 43 Resp: 10112


Ion Ratio Lower Upper

43 100

70 32.2 38.6 58.0#

55 43.8 22.2 33.2#

61 0.0 21.3 31.9#






Abundance Ion 43.00 (42.70 to 43.70): VW

Ion 70.00 (69.70 to 70.70): VW

Ion 55.00 (54.70 to 55.70): VW

Ion 61.00 (60.70 to 61.70): VW





#75

1,1,2,2-Tetrachloroethane

Concen: 5.356 ug/l

RT: 12.69 min Scan# 1769

Delta R.T. -0.03 min

Lab File: VW016282.D

Acq: 21 Aug 2020 00:45

Instrument :

MSVOA_W

ClientSampled :

CHRT-26982RE





Tot Ion: 83 Resp: 4357

Ion Ratio Lower Upper

83 100

131 3.5 5.6 16.8#

85 40.9 31.3 93.9



#76

1,2,3-Trichloropropane

Concen: 78.415 ug/l

RT: 12.62 min Scan# 1758

Delta R.T. -0.15 min

Lab File: VW016282.D




Acq: 21 Aug 2020 00:45


Tgt Ion: 75 Resp: 46011

Ion Ratio Lower Upper

75 100

77 3.8 173.6 520.6#









#80
 1,3,5-Trimethylbenzene
 Concen: 2.695 ug/l
 RT: 12.94 min Scan# 1811
 Delta R.T. -0.00 min
 Lab File: VW016282.D
 Acq: 21 Aug 2020 00:45




Instrument :
 MSVOA_W
 ClientSampleId :
 CHRT-26982RE


Tot Ion:105 Resp: 11246
 Ion Ratio Lower Upper
 105 100
 120 45.6 25.1 75.2



#81
 trans-1,4-Dichloro-2-butene
 Concen: 188.090 ug/l
 RT: 12.62 min Scan# 1758
 Delta R.T. 0.10 min
 Lab File: VW016282.D
 Acq: 21 Aug 2020 00:45

Tgt Ion: 75 Resp: 46011
 Ion Ratio Lower Upper
 75 100
 53 2.9 74.6 111.8#
 89 0.2 37.2 55.8#









#83
tert-Butylbenzene
Concen: 1.928 ug/l
RT: 13.44 min Scan# 1893
Delta R.T. 0.24 min
Lab File: VW016282.D
Acq: 21 Aug 2020 00:45




Instrument :
MSVOA_W
ClientSampleId :
CHRT-26982RE


Tot Ion:119 Resp: 6596
Ion Ratio Lower Upper
119 100
91 52.7 33.3 99.9
134 31.1 13.5 40.5



#84
1,2,4-Trimethylbenzene
Concen: 5.157 ug/l
RT: 13.26 min Scan# 1862
Delta R.T. 0.01 min
Lab File: VW016282.D
Acq: 21 Aug 2020 00:45

Tgt Ion:105 Resp: 21299
Ion Ratio Lower Upper
105 100
120 46.6 22.2 66.6









#86
p-Isopropyltoluene
Concen: 6.108 ug/l
RT: 13.50 min Scan# 1902
Delta R.T. -0.00 min
Lab File: VW016282.D
Acq: 21 Aug 2020 00:45




Instrument :
MSVOA_W
ClientSampleId :
CHRT-26982RE


Tot Ion:119 Resp: 27462
Ion Ratio Lower Upper
119 100
134 27.3 12.8 38.4
91 25.3 11.8 35.3



#89
n-Butylbenzene
Concen: 2.302 ug/l
RT: 13.79 min Scan# 1950
Delta R.T. -0.03 min
Lab File: VW016282.D
Acq: 21 Aug 2020 00:45

Tgt Ion: 91 Resp: 9576
Ion Ratio Lower Upper
91 100
92 0.0 27.4 82.2#
134 121.7 13.0 39.0#





#90

Hexachloroethane

Concen: 3.376 ug/l

RT: 14.10 min Scan# 2000

Delta R.T. -0.00 min

Lab File: VW016282.D

Acq: 21 Aug 2020 00:45

Instrument :

MSVOA_W

ClientSampled :





CHRT-26982RE

Tot Ion:117 Resp: 2833

Ion Ratio Lower Upper

117 100

201 18.7 33.8 101.3#



#92

1,2-Dibromo-3-Chloropropane

Concen: 2.903 ug/l

RT: 14.50 min Scan# 2066

Delta R.T. 0.01 min

Lab File: VW016282.D

Acq: 21 Aug 2020 00:45




Tgt Ion: 75 Resp: 363


Ion Ratio Lower Upper

75 100

155 34.7 41.3 124.1#

157 0.0 53.0 159.0#









#93
 1,2,4-Trichlorobenzene
 Concen: 1.323 ug/l
 RT: 15.13 min Scan# 2170
 Delta R.T. -0.00 min
 Lab File: VW016282.D
 Acq: 21 Aug 2020 00:45




Instrument :
 MSVOA_W
 ClientSampleId :
 CHRT-26982RE


Tot Ion:180 Resp: 1481
 Ion Ratio Lower Upper
 180 100
 182 304.6 46.0 138.0#
 145 2965.3 16.9 50.7#



#95
 Naphthalene
 Concen: 5.945 ug/l
 RT: 15.36 min Scan# 2208
 Delta R.T. -0.01 min
 Lab File: VW016282.D
 Acq: 21 Aug 2020 00:45

Tgt Ion:128 Resp: 5987
 Ion Ratio Lower Upper
 128 100
 127 0.0 10.2 15.4#
 129 10.8 8.7 13.1





#96
 1,2,3-Trichlorobenzene
 Concen: 8.799 ug/l
 RT: 15.58 min Scan# 2243
 Delta R.T. 0.02 min
 Lab File: VW016282.D
 Acq: 21 Aug 2020 00:45

Instrument :
 MSVOA_W
 ClientSampleId :
 CHRT-26982RE

Tot Ion	Ratio	Lower	Upper
180	100		
182	0.0	47.7	143.1#
145	0.0	17.4	52.3#

