

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW090120\
 Data File : VW016432.D
 Acq On : 01 Sep 2020 15:18
 Operator : SY/VA
 Sample : TEST 1
 Misc : 5.00G/5ML/MSVOA W/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 TEST 1

Quant Time: Sep 02 09:07:50 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W082820S.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 28 15:01:43 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	128385	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	222926	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	203432	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	91741	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	76066	59.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	118.38%	
35) Dibromofluoromethane	7.88	113	71862	50.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.86%	
50) Toluene-d8	10.33	98	262468	49.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.34%	
62) 4-Bromofluorobenzene	12.62	95	94853	47.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.74%	
Target Compounds						
16) Acetone	4.15	43	2016	9.238	ug/l #	85
29) Tetrahydrofuran	7.55	42	1770	10.013	ug/l #	73

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW090120\
Data File : VW016432.D
Acq On : 01 Sep 2020 15:18
Operator : SY/VA
Sample : TEST 1
Misc : 5.00G/5ML/MSVOA W/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
TEST 1

Quant Time: Sep 02 09:07:50 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W082820S.M
Quant Title : SW846 8260
QLast Update : Fri Aug 28 15:01:43 2020
Response via : Initial Calibration

