

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW090419\
 Data File : VW012593.D
 Acq On : 04 Sep 2019 14:32
 Operator : SY/VA
 Sample : K4619-21
 Misc : 5.53G/5ML/MSVOA W/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 TP7-COMP

Quant Time: Sep 04 15:04:52 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W082319S.M
 Quant Title : SW846 8260
 QLast Update : Mon Aug 26 07:18:31 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	224755	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	361127	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	288798	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	113364	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.30	65	129781	48.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.14%	
35) Dibromofluoromethane	7.88	113	107514	49.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.76%	
50) Toluene-d8	10.32	98	434867	49.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.50%	
62) 4-Bromofluorobenzene	12.62	95	132549	43.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.04%	
Target Compounds						
16) Acetone	4.14	43	38851	56.886	ug/l	99
20) Methylene Chloride	4.91	84	15541	6.019	ug/l	91
86) p-Isopropyltoluene	13.50	119	51388	6.680	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW090419\
Data File : VW012593.D
Acq On : 04 Sep 2019 14:32
Operator : SY/VA
Sample : K4619-21
Misc : 5.53G/5ML/MSVOA W/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
TP7-COMP

Quant Time: Sep 04 15:04:52 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W082319S.M
Quant Title : SW846 8260
QLast Update : Mon Aug 26 07:18:31 2019
Response via : Initial Calibration

