

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW090419\
 Data File : VW012594.D
 Acq On : 04 Sep 2019 14:58
 Operator : SY/VA
 Sample : K4619-24
 Misc : 5.84G/5ML/MSVOA W/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 TP8-COMP

Quant Time: Sep 04 15:25:31 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W082319S.M
 Quant Title : SW846 8260
 QLast Update : Mon Aug 26 07:18:31 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	245147	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	399835	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	337296	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	138324	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	165146	56.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	113.34%	
35) Dibromofluoromethane	7.88	113	125052	52.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.80%	
50) Toluene-d8	10.32	98	489652	50.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.20%	
62) 4-Bromofluorobenzene	12.62	95	157534	46.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.36%	
Target Compounds						
16) Acetone	4.14	43	52827	70.915	ug/l	96
20) Methylene Chloride	4.91	84	12017	4.267	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW090419\
Data File : VW012594.D
Acq On : 04 Sep 2019 14:58
Operator : SY/VA
Sample : K4619-24
Misc : 5.84G/5ML/MSVOA W/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
TP8-COMP

Quant Time: Sep 04 15:25:31 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W082319S.M
Quant Title : SW846 8260
QLast Update : Mon Aug 26 07:18:31 2019
Response via : Initial Calibration

