

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW090419\
 Data File : VW012605.D
 Acq On : 04 Sep 2019 21:30
 Operator : SY/VA
 Sample : K4613-06
 Misc : 6.57G/5ML/MSVOA W/SOIL
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 TP-10-D2

Quant Time: Sep 05 07:20:46 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W082319S.M
 Quant Title : SW846 8260
 QLast Update : Mon Aug 26 07:18:31 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	248421	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	408855	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	346860	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	140665	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.30	65	166102	56.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	112.50%	
35) Dibromofluoromethane	7.88	113	129168	52.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.86%	
50) Toluene-d8	10.32	98	502675	50.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.60%	
62) 4-Bromofluorobenzene	12.61	95	160813	46.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.20%	
Target Compounds						
20) Methylene Chloride	4.92	84	17035	5.969	ug/l	Qvalue 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW090419\
Data File : VW012605.D
Acq On : 04 Sep 2019 21:30
Operator : SY/VA
Sample : K4613-06
Misc : 6.57G/5ML/MSVOA W/SOIL
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
TP-10-D2

Quant Time: Sep 05 07:20:46 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W082319S.M
Quant Title : SW846 8260
QLast Update : Mon Aug 26 07:18:31 2019
Response via : Initial Calibration

