

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW090419\
 Data File : VW012606.D
 Acq On : 04 Sep 2019 21:56
 Operator : SY/VA
 Sample : K4601-02
 Misc : 4.25G/5ML/MSVOA W/SOIL
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 T12-E1-(0)

Quant Time: Sep 05 07:23:15 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W082319S.M
 Quant Title : SW846 8260
 QLast Update : Mon Aug 26 07:18:31 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	241719	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	396749	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	322092	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	114824	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.30	65	162729	56.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	113.26%	
35) Dibromofluoromethane	7.88	113	123906	52.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.64%	
50) Toluene-d8	10.32	98	479167	49.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.80%	
62) 4-Bromofluorobenzene	12.62	95	141708	41.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	83.74%	
Target Compounds						
20) Methylene Chloride	4.92	84	17165	6.182	ug/l	95
29) Tetrahydrofuran	7.53	42	5167	8.628	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW090419\
Data File : VW012606.D
Acq On : 04 Sep 2019 21:56
Operator : SY/VA
Sample : K4601-02
Misc : 4.25G/5ML/MSVOA W/SOIL
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
T12-E1-(0)

Quant Time: Sep 05 07:23:15 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W082319S.M
Quant Title : SW846 8260
QLast Update : Mon Aug 26 07:18:31 2019
Response via : Initial Calibration

