

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW090519\
 Data File : VW012636.D
 Acq On : 05 Sep 2019 13:07
 Operator : SY/VA
 Sample : K4593-01RE
 Misc : 5.16G/5ML/MSVOA W/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 S-3-(5-7)RE

Quant Time: Sep 05 16:54:24 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W082319S.M
 Quant Title : SW846 8260
 QLast Update : Mon Aug 26 07:18:31 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	234077	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	363393	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	258052	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	69897	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	141523	50.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.72%	
35) Dibromofluoromethane	7.88	113	114316	52.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.40%	
50) Toluene-d8	10.32	98	439833	50.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.02%	
62) 4-Bromofluorobenzene	12.62	95	103189	33.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	66.56%	
Target Compounds						
16) Acetone	4.14	43	5927	8.333	ug/l	95
64) Tetrachloroethene	10.87	164	2689	1.607	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW090519\
Data File : VW012636.D
Acq On : 05 Sep 2019 13:07
Operator : SY/VA
Sample : K4593-01RE
Misc : 5.16G/5ML/MSVOA W/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
S-3-(5-7)RE

Quant Time: Sep 05 16:54:24 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W082319S.M
Quant Title : SW846 8260
QLast Update : Mon Aug 26 07:18:31 2019
Response via : Initial Calibration

