

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW090519\
 Data File : VW012643.D
 Acq On : 05 Sep 2019 16:20
 Operator : SY/VA
 Sample : K4639-03RE
 Misc : 5.10G/5ML/MSVOA W/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 DW-01RE

Quant Time: Sep 05 17:09:04 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W082319S.M
 Quant Title : SW846 8260
 QLast Update : Mon Aug 26 07:18:31 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	225470	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	374879	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	236397	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	49876	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	148079	55.25	ug/l	0.00
Spiked Amount	50.000		Recovery	=	110.50%	
35) Dibromofluoromethane	7.88	113	119336	53.33	ug/l	0.00
Spiked Amount	50.000		Recovery	=	106.66%	
50) Toluene-d8	10.32	98	401292	44.23	ug/l	0.00
Spiked Amount	50.000		Recovery	=	88.46%	
62) 4-Bromofluorobenzene	12.62	95	84937	26.56	ug/l	0.00
Spiked Amount	50.000		Recovery	=	53.12%	
Target Compounds						
16) Acetone	4.13	43	82574	120.522	ug/l	92
20) Methylene Chloride	4.92	84	9702	3.746	ug/l	97
25) 2-Butanone	7.17	43	36155	38.834	ug/l	98
51) 4-Methyl-2-Pentanone	10.21	43	21365	10.682	ug/l	99
67) Ethyl Benzene	11.73	91	13457	1.501	ug/l	93
86) p-Isopropyltoluene	13.50	119	4421	1.306	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW090519\
Data File : VW012643.D
Acq On : 05 Sep 2019 16:20
Operator : SY/VA
Sample : K4639-03RE
Misc : 5.10G/5ML/MSVOA W/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
DW-01RE

Quant Time: Sep 05 17:09:04 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W082319S.M
Quant Title : SW846 8260
QLast Update : Mon Aug 26 07:18:31 2019
Response via : Initial Calibration

