

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW090519\
 Data File : VW012644.D
 Acq On : 05 Sep 2019 16:45
 Operator : SY/VA
 Sample : K4639-04RE
 Misc : 5.06G/5ML/MSVOA W/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 DW-02RE

Quant Time: Sep 05 17:16:52 2019
 Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W082319S.M
 Quant Title : SW846 8260
 QLast Update : Mon Aug 26 07:18:31 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	163606	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	272219	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	146466	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	27288	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.30	65	118616	60.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	121.98%	
35) Dibromofluoromethane	7.88	113	88461	54.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.88%	
50) Toluene-d8	10.32	98	269331	40.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	81.76%	
62) 4-Bromofluorobenzene	12.62	95	51548	22.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	44.40%	
Target Compounds						
					Qvalue	
16) Acetone	4.14	43	68266	137.314	ug/l	96
17) Carbon Disulfide	4.37	76	34217	6.899	ug/l	97
20) Methylene Chloride	4.91	84	8140	4.331	ug/l	89
25) 2-Butanone	7.18	43	25850	38.264	ug/l	95
51) 4-Methyl-2-Pentanone	10.21	43	14414	9.925	ug/l	99
52) Toluene	10.38	92	7186	1.557	ug/l	96
84) 1,2,4-Trimethylbenzene	13.26	105	7266	4.234	ug/l	99
85) sec-Butylbenzene	13.38	105	4549	2.194	ug/l #	69

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW090519\
Data File : VW012644.D
Acq On : 05 Sep 2019 16:45
Operator : SY/VA
Sample : K4639-04RE
Misc : 5.06G/5ML/MSVOA W/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
DW-02RE

Quant Time: Sep 05 17:16:52 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W082319S.M
Quant Title : SW846 8260
QLast Update : Mon Aug 26 07:18:31 2019
Response via : Initial Calibration

