

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW090519\
 Data File : VW012653.D
 Acq On : 05 Sep 2019 20:36
 Operator : SY/VA
 Sample : K4579-04
 Misc : 4.09G/5ML/MSVOA W/SOIL
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 PAR-12-A4-(0)

Quant Time: Sep 06 08:24:57 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W082319S.M
 Quant Title : SW846 8260
 QLast Update : Mon Aug 26 07:18:31 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	257781	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	415607	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	341338	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.55	152	132271	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	161503	52.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.40%	
35) Dibromofluoromethane	7.88	113	127166	51.26	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.52%	
50) Toluene-d8	10.32	98	501459	49.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.70%	
62) 4-Bromofluorobenzene	12.62	95	155267	43.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.58%	
Target Compounds						
16) Acetone	4.15	43	59824	76.372	ug/l	94
20) Methylene Chloride	4.92	84	27312	9.223	ug/l	95
29) Tetrahydrofuran	7.54	42	5545	8.683	ug/l #	68
52) Toluene	10.38	92	13282	1.885	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW090519\
Data File : VW012653.D
Acq On : 05 Sep 2019 20:36
Operator : SY/VA
Sample : K4579-04
Misc : 4.09G/5ML/MSVOA W/SOIL
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
PAR-12-A4-(0)

Quant Time: Sep 06 08:24:57 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W082319S.M
Quant Title : SW846 8260
QLast Update : Mon Aug 26 07:18:31 2019
Response via : Initial Calibration

