

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW090519\
 Data File : VW012666.D
 Acq On : 06 Sep 2019 03:04
 Operator : SY/VA
 Sample : K4579-11
 Misc : 4.24G/5ML/MSVOA W/SOIL
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 PAR-12-B3-(0)

Quant Time: Sep 06 08:42:55 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W082319S.M
 Quant Title : SW846 8260
 QLast Update : Mon Aug 26 07:18:31 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	239329	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	417464	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	323963	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	120038	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	169564	59.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	119.20%	
35) Dibromofluoromethane	7.88	113	119229	47.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.70%	
50) Toluene-d8	10.32	98	531347	52.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.18%	
62) 4-Bromofluorobenzene	12.62	95	145075	40.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	81.46%	
Target Compounds						
16) Acetone	4.14	43	44654	61.401	ug/l	Qvalue 100
20) Methylene Chloride	4.91	84	18217	6.626	ug/l	93
29) Tetrahydrofuran	7.53	42	5491	9.261	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW090519\
Data File : VW012666.D
Acq On : 06 Sep 2019 03:04
Operator : SY/VA
Sample : K4579-11
Misc : 4.24G/5ML/MSVOA W/SOIL
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
PAR-12-B3-(0)

Quant Time: Sep 06 08:42:55 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W082319S.M
Quant Title : SW846 8260
QLast Update : Mon Aug 26 07:18:31 2019
Response via : Initial Calibration

