

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW090519\
 Data File : VW012670.D
 Acq On : 06 Sep 2019 04:47
 Operator : SY/VA
 Sample : K4579-15
 Misc : 6.16G/5ML/MSVOA W/SOIL
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 PAR-12-B3-(3.5)

Quant Time: Sep 06 08:48:35 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W082319S.M
 Quant Title : SW846 8260
 QLast Update : Mon Aug 26 07:18:31 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	258182	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	409978	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	335600	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	120085	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.30	65	165636	53.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.94%	
35) Dibromofluoromethane	7.88	113	125517	51.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.58%	
50) Toluene-d8	10.32	98	498973	50.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.58%	
62) 4-Bromofluorobenzene	12.62	95	148104	42.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	84.68%	
Target Compounds						
16) Acetone	4.14	43	38308	48.829	ug/l	97
20) Methylene Chloride	4.90	84	16475	5.555	ug/l	91
29) Tetrahydrofuran	7.53	42	4709	7.362	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW090519\
Data File : VW012670.D
Acq On : 06 Sep 2019 04:47
Operator : SY/VA
Sample : K4579-15
Misc : 6.16G/5ML/MSVOA W/SOIL
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
PAR-12-B3-(3.5)

Quant Time: Sep 06 08:48:35 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W082319S.M
Quant Title : SW846 8260
QLast Update : Mon Aug 26 07:18:31 2019
Response via : Initial Calibration

