

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW090519\
 Data File : VW012671.D
 Acq On : 06 Sep 2019 05:13
 Operator : SY/VA
 Sample : K4579-16
 Misc : 3.60G/5ML/MSVOA W/SOIL
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 PAR-12-B4-(3.5)

Quant Time: Sep 06 08:49:54 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W082319S.M
 Quant Title : SW846 8260
 QLast Update : Mon Aug 26 07:18:31 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	254695	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	427083	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	356279	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	124270	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.30	65	168369	55.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.22%	
35) Dibromofluoromethane	7.88	113	124401	48.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.60%	
50) Toluene-d8	10.32	98	528924	51.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.34%	
62) 4-Bromofluorobenzene	12.62	95	153264	42.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	84.12%	
Target Compounds						
16) Acetone	4.15	43	6383	8.247	ug/l	94
20) Methylene Chloride	4.92	84	22764	7.780	ug/l #	91
29) Tetrahydrofuran	7.54	42	5426	8.599	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW090519\
Data File : VW012671.D
Acq On : 06 Sep 2019 05:13
Operator : SY/VA
Sample : K4579-16
Misc : 3.60G/5ML/MSVOA W/SOIL
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
PAR-12-B4-(3.5)

Quant Time: Sep 06 08:49:54 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W082319S.M
Quant Title : SW846 8260
QLast Update : Mon Aug 26 07:18:31 2019
Response via : Initial Calibration

