

Data File : VW012693.D
Acq On : 06 Sep 2019 15:59
Operator : SY/VA
Sample : K4580-07
Misc : 2.08G/5ML/MSVOA_W/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
T12-B4-(0)

Quant Time: Sep 07 05:48:23 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W082319S.M
Quant Title : SW846 8260
QLast Update : Mon Aug 26 07:18:31 2019
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
-----	-----	-----	-----	-----	-----	-----
1) Pentafluorobenzene	7.95	168	254168	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	397269	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	283126	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.55	152	82803	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	150289	49.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.48%	
35) Dibromofluoromethane	7.88	113	115222	48.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.18%	
50) Toluene-d8	10.32	98	450133	46.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.64%	
62) 4-Bromofluorobenzene	12.62	95	157137	46.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.72%	

Target Compounds	Qvalue
-----	-----

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VW012693.D
Acq On : 06 Sep 2019 15:59
Operator : SY/VA
Sample : K4580-07
Misc : 2.08G/5ML/MSVOA_W/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
T12-B4-(0)

Quant Time: Sep 07 05:48:23 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W082319S.M
Quant Title : SW846 8260
QLast Update : Mon Aug 26 07:18:31 2019
Response via : Initial Calibration

