

Data Path : Z:\voasrv\HPCHEM1\MSVOA W\Data\VW092519\
 Data File : VW013296.D
 Acq On : 25 Sep 2019 17:11
 Operator : SY/VA
 Sample : VSTDCCC050
 Misc : 5.00G/5ML/MSVOA W/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_W
 LabSampleId :
 VSTDCCC050

Quant Time: Sep 25 17:46:07 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W092019S.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 20 15:58:08 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	94	0.00
2 T	Dichlorodifluoromethane	0.288	0.286	0.7	100	0.00
3 P	Chloromethane	0.371	0.313	15.6	87	0.00
4 C	Vinyl Chloride	0.478	0.429	10.3#	90	0.00
5 T	Bromomethane	0.303	0.255	15.8	81	-0.01
6 T	Chloroethane	0.282	0.255	9.6	87	-0.01
7 T	Trichlorofluoromethane	0.276	0.236	14.5	83	0.00
8 T	Diethyl Ether	0.235	0.230	2.1	93	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.449	0.407	9.4	87	0.00
10 T	Methyl Iodide	0.696	0.652	6.3	89	0.00
11 T	Tert butyl alcohol	0.034	0.031	8.8	109	0.00
12 CM	1,1-Dichloroethene	0.465	0.428	8.0#	89	0.00
13 T	Acrolein	0.023	0.020	13.0	75	0.00
14 T	Allyl chloride	0.740	0.708	4.3	90	0.00
15 T	Acrylonitrile	0.102	0.105	-2.9	100	0.00
16 T	Acetone	0.092	0.088	4.3	96	-0.01
17 T	Carbon Disulfide	1.343	1.197	10.9	84	0.00
18 T	Methyl Acetate	0.259	0.300	-15.8	118	0.00
19 T	Methyl tert-butyl Ether	0.712	0.691	2.9	92	0.00
20 T	Methylene Chloride	0.521	0.462	11.3	91	0.00
21 T	trans-1,2-Dichloroethene	0.502	0.473	5.8	89	0.00
22 T	Diisopropyl ether	1.406	1.367	2.8	90	0.00
23 T	Vinyl Acetate	0.834	0.804	3.6	90	0.00
24 P	1,1-Dichloroethane	0.849	0.798	6.0	89	0.00
25 T	2-Butanone	0.138	0.134	2.9	98	0.00
26 T	2,2-Dichloropropane	0.563	0.460	18.3	82	0.00
27 T	cis-1,2-Dichloroethene	0.530	0.510	3.8	90	0.00
28 T	Bromochloromethane	0.326	0.345	-5.8	92	0.00
29 T	Tetrahydrofuran	0.085	0.087	-2.4	101	0.00
30 C	Chloroform	0.831	0.775	6.7#	91	0.00
31 T	Cyclohexane	0.885	0.779	12.0	87	0.00
32 T	1,1,1-Trichloroethane	0.673	0.631	6.2	89	0.00
33 S	1,2-Dichloroethane-d4	0.408	0.405	0.7	93	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	95	0.00
35 S	Dibromofluoromethane	0.275	0.281	-2.2	96	0.00
36 T	1,1-Dichloropropene	0.484	0.452	6.6	90	0.00
37 T	Ethyl Acetate	0.206	0.203	1.5	97	0.00
38 T	Carbon Tetrachloride	0.434	0.406	6.5	88	0.00
39 T	Methylcyclohexane	0.622	0.564	9.3	85	0.00
40 TM	Benzene	1.351	1.287	4.7	91	0.00
41 T	Methacrylonitrile	0.123	0.129	-4.9	100	0.00
42 TM	1,2-Dichloroethane	0.363	0.343	5.5	90	0.00
43 T	Isopropyl Acetate	0.394	0.384	2.5	94	0.00
44 TM	Trichloroethene	0.379	0.352	7.1	89	0.00
45 C	1,2-Dichloropropane	0.330	0.313	5.2#	90	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA W\Data\VW092519\
 Data File : VW013296.D
 Acq On : 25 Sep 2019 17:11
 Operator : SY/VA
 Sample : VSTDCCC050
 Misc : 5.00G/5ML/MSVOA W/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_W
 LabSampleId :
 VSTDCCC050

Quant Time: Sep 25 17:46:07 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W092019S.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 20 15:58:08 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.160	0.157	1.9	94	0.00
47 T	Bromodichloromethane	0.407	0.390	4.2	90	0.00
48 T	Methyl methacrylate	0.185	0.197	-6.5	103	0.00
49 T	1,4-Dioxane	0.002	0.003	-50.0#	116	-0.01
50 S	Toluene-d8	1.151	1.156	-0.4	93	0.00
51 T	4-Methyl-2-Pentanone	0.198	0.198	0.0	97	0.00
52 CM	Toluene	0.866	0.820	5.3#	89	0.00
53 T	t-1,3-Dichloropropene	0.419	0.407	2.9	90	0.00
54 T	cis-1,3-Dichloropropene	0.509	0.492	3.3	91	0.00
55 T	1,1,2-Trichloroethane	0.239	0.232	2.9	93	0.00
56 T	Ethyl methacrylate	0.313	0.322	-2.9	95	0.00
57 T	1,3-Dichloropropane	0.418	0.409	2.2	93	0.00
58 T	2-Chloroethyl Vinyl ether	0.145	0.155	-6.9	96	0.00
59 T	2-Hexanone	0.136	0.137	-0.7	97	0.00
60 T	Dibromochloromethane	0.272	0.268	1.5	92	0.00
61 T	1,2-Dibromoethane	0.228	0.223	2.2	95	0.00
62 S	4-Bromofluorobenzene	0.393	0.403	-2.5	96	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	96	0.00
64 T	Tetrachloroethene	0.379	0.354	6.6	90	0.00
65 PM	Chlorobenzene	1.044	0.988	5.4	92	0.00
66 T	1,1,1,2-Tetrachloroethane	0.357	0.345	3.4	91	0.00
67 C	Ethyl Benzene	1.892	1.778	6.0#	89	0.00
68 T	m/p-Xylenes	0.721	0.682	5.4	89	0.00
69 T	o-Xylene	0.670	0.629	6.1	89	0.00
70 T	Styrene	1.151	1.120	2.7	90	0.00
71 P	Bromoform	0.188	0.183	2.7	92	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	97	0.00
73 T	Isopropylbenzene	3.694	3.435	7.0	89	0.00
74 T	N-amyl acetate	0.820	0.781	4.8	91	0.00
75 P	1,1,2,2-Tetrachloroethane	0.618	0.592	4.2	95	0.00
76 T	1,2,3-Trichloropropane	0.442	0.464	-5.0	96	0.00
77 T	Bromobenzene	0.875	0.822	6.1	89	0.00
78 T	n-propylbenzene	4.323	3.978	8.0	87	0.00
79 T	2-Chlorotoluene	2.422	2.245	7.3	89	0.00
80 T	1,3,5-Trimethylbenzene	3.105	2.871	7.5	88	0.00
81 T	trans-1,4-Dichloro-2-butene	0.197	0.197	0.0	96	0.00
82 T	4-Chlorotoluene	2.554	2.337	8.5	89	0.00
83 T	tert-Butylbenzene	2.725	2.539	6.8	88	0.00
84 T	1,2,4-Trimethylbenzene	3.090	2.854	7.6	88	0.00
85 T	sec-Butylbenzene	3.757	3.439	8.5	86	0.00
86 T	p-Isopropyltoluene	3.485	3.212	7.8	87	0.00
87 T	1,3-Dichlorobenzene	1.685	1.542	8.5	88	0.00
88 T	1,4-Dichlorobenzene	1.652	1.530	7.4	89	0.00
89 T	n-Butylbenzene	3.182	2.906	8.7	86	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA W\Data\VW092519\
Data File : VW013296.D
Acq On : 25 Sep 2019 17:11
Operator : SY/VA
Sample : VSTDCCC050
Misc : 5.00G/5ML/MSVOA W/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_W
LabSampleId :
VSTDCCC050

Quant Time: Sep 25 17:46:07 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W092019S.M
Quant Title : SW846 8260
QLast Update : Fri Sep 20 15:58:08 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.585	0.541	7.5	87	0.00
91 T	1,2-Dichlorobenzene	1.457	1.376	5.6	91	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.094	0.094	0.0	97	0.00
93 T	1,2,4-Trichlorobenzene	1.044	0.990	5.2	89	0.00
94 T	Hexachlorobutadiene	0.714	0.656	8.1	89	0.00
95 T	Naphthalene	1.708	1.763	-3.2	95	0.00
96 T	1,2,3-Trichlorobenzene	0.902	0.872	3.3	92	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6