

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW101321\
 Data File : VW020464.D
 Acq On : 13 Oct 2021 16:39
 Operator : SY/VA
 Sample : M4181-06
 Misc : 5.29g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 C0B20

Quant Time: Oct 14 01:03:21 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM100821SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Oct 14 00:58:33 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.847	114	280857	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.634	117	244949	25.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.560	152	97603	25.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	83238	20.37	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	81.48%
7) Chloroethane-d5	2.879	69	48206	22.33	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	89.32%
11) 1,1-Dichloroethene-d2	4.013	63	120978	14.45	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	57.80%
21) 2-Butanone-d5	7.086	46	35153	28.26	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	56.52%
24) Chloroform-d	7.653	84	175555	19.61	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	78.44%
26) 1,2-Dichloroethane-d4	8.305	65	91347	17.87	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	71.48%
32) Benzene-d6	8.274	84	339440	21.07	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	84.28%
36) 1,2-Dichloropropane-d6	9.274	67	109274	20.81	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	83.24%
41) Toluene-d8	10.323	98	295523	20.44	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	81.76%
43) trans-1,3-Dichloroprop...	10.579	79	35738	16.75	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	67.00%
47) 2-Hexanone-d5	10.926	63	21815	27.00	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	54.00%
56) 1,1,2,2-Tetrachloroeth...	12.694	84	62727	15.80	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	63.20%
66) 1,2-Dichlorobenzene-d4	13.853	152	77612	20.33	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	81.32%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW101321\
Data File : VW020464.D
Acq On : 13 Oct 2021 16:39
Operator : SY/VA
Sample : M4181-06
Misc : 5.29g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
C0B20

Quant Time: Oct 14 01:03:21 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM100821SMA.M
Quant Title : SFAM01.0
QLast Update : Thu Oct 14 00:58:33 2021
Response via : Initial Calibration

