

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW102021\
 Data File : VW020649.D
 Acq On : 20 Oct 2021 13:00
 Operator : SY/VA
 Sample : M4264-04
 Misc : 4.23g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 GB7G9

Quant Time: Oct 21 01:11:40 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM100821SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Oct 21 01:08:32 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	211490	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.634	117	193406	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.560	152	74197	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	60674	19.716	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	78.880%
7) Chloroethane-d5	2.879	69	22502	13.843	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	55.360%
11) 1,1-Dichloroethene-d2	4.013	63	85799	13.606	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	54.440%
21) 2-Butanone-d5	7.080	46	40640	43.393	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	86.780%
24) Chloroform-d	7.647	84	139735	20.728	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	82.920%
26) 1,2-Dichloroethane-d4	8.305	65	87457	22.714	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	90.840%
32) Benzene-d6	8.275	84	272034	21.383	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	85.520%
36) 1,2-Dichloropropane-d6	9.274	67	91493	22.066	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	88.280%
41) Toluene-d8	10.323	98	244569	21.426	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	85.720%
43) trans-1,3-Dichloroprop...	10.579	79	36939	21.924	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	87.680%
47) 2-Hexanone-d5	10.927	63	31430	49.276	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	98.560%
56) 1,1,2,2-Tetrachloroeth...	12.688	84	67491	21.535	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	86.120%
66) 1,2-Dichlorobenzene-d4	13.853	152	65369	22.522	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	90.080%
Target Compounds					Qvalue	
13) Acetone	4.135	43	10834	12.600	ug/L	73

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW102021\
Data File : VW020649.D
Acq On : 20 Oct 2021 13:00
Operator : SY/VA
Sample : M4264-04
Misc : 4.23g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
GB7G9

Quant Time: Oct 21 01:11:40 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM100821SMA.M
Quant Title : SFAM01.0
QLast Update : Thu Oct 21 01:08:32 2021
Response via : Initial Calibration

