

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW102021\
 Data File : VW020650.D
 Acq On : 20 Oct 2021 13:24
 Operator : SY/VA
 Sample : M4264-05
 Misc : 4.01g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 GB7H0

Quant Time: Oct 21 01:11:50 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM100821SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Oct 21 01:08:32 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.848	114	216153	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.634	117	199306	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.560	152	77848	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	63474	20.181	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	80.720%
7) Chloroethane-d5	2.879	69	21001	12.640	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	50.560%
11) 1,1-Dichloroethene-d2	4.013	63	89522	13.890	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	55.560%
21) 2-Butanone-d5	7.080	46	48098	50.248	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	100.500%
24) Chloroform-d	7.653	84	148419	21.541	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	86.160%
26) 1,2-Dichloroethane-d4	8.305	65	92600	23.531	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	94.120%
32) Benzene-d6	8.275	84	288324	21.993	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	87.960%
36) 1,2-Dichloropropane-d6	9.274	67	96786	22.652	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	90.600%
41) Toluene-d8	10.323	98	255913	21.756	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	87.040%
43) trans-1,3-Dichloroprop...	10.579	79	39606	22.811	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	91.240%
47) 2-Hexanone-d5	10.927	63	35909	54.632	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	109.260%
56) 1,1,2,2-Tetrachloroeth...	12.695	84	73784	22.846	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	91.400%
66) 1,2-Dichlorobenzene-d4	13.853	152	70479	23.144	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	92.560%
Target Compounds					Qvalue	
13) Acetone	4.123	43	15745	17.916	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW102021\
Data File : VW020650.D
Acq On : 20 Oct 2021 13:24
Operator : SY/VA
Sample : M4264-05
Misc : 4.01g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
GB7H0

Quant Time: Oct 21 01:11:50 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM100821SMA.M
Quant Title : SFAM01.0
QLast Update : Thu Oct 21 01:08:32 2021
Response via : Initial Calibration

