

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW110420\
 Data File : VW017082.D
 Acq On : 04 Nov 2020 19:37
 Operator : SY/VA
 Sample : VIBLK03
 Misc : 5.00G/10ML/MSVOA W/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VIBLK03

Quant Time: Nov 05 07:18:34 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM110420S.M
 Quant Title : VOC Analysis
 QLast Update : Wed Nov 04 16:12:16 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.84	114	496409	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.63	117	463381	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.56	152	221519	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.35	65	172313	27.56	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	110.24%
7) Chloroethane-d5	2.89	69	126520	26.26	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	105.04%
10) 1,1-Dichloroethene-d2	4.02	63	219011	19.58	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	78.32%
20) 2-Butanone-d5	7.08	46	82808	59.22	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	118.44%
24) Chloroform-d	7.65	84	305460	26.36	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	105.44%
26) 1,2-Dichloroethane-d4	8.31	65	160312	27.83	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	111.32%
29) Benzene-d6	8.27	84	626831	26.56	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	106.24%
33) 1,2-Dichloropropane-d6	9.27	67	174820	26.78	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	107.12%
37) Toluene-d8	10.32	98	587251	26.39	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	105.56%
38) trans-1,3-Dichloropropene-	10.58	79	76877	26.12	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	104.48%
39) 2-Hexanone-d5	10.93	63	68678	60.29	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	120.58%
48) 1,1,2,2-Tetrachloroethane-	12.69	84	161448	28.40	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	113.60%
61) 1,2-Dichlorobenzene-d4	13.85	152	218488	28.00	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	112.00%

Target Compounds					Ovalue
16) Methylene chloride	4.92	84	9459	1.413 ug/L	94
43) 4-Methyl-2-pentanone	10.21	43	5721	1.630 ug/L	98
47) Tetrachloroethene	10.86	164	2305	0.375 ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW110420\
Data File : VW017082.D
Acq On : 04 Nov 2020 19:37
Operator : SY/VA
Sample : VIBLK03
Misc : 5.00G/10ML/MSVOA W/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VIBLK03

Quant Time: Nov 05 07:18:34 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM110420S.M
Quant Title : VOC Analysis
QLast Update : Wed Nov 04 16:12:16 2020
Response via : Initial Calibration

