

Quantitation Report (Qedit)


Data Path : Z:\voasrv\HPCHEM1\MSVOA W\Data\VW110420\
 Data File : VW017080.D
 Acq On : 04 Nov 2020 18:04
 Operator : SY/VA
 Sample : VW1104SBL01
 Misc : 5.00G/10ML/MSVOA W/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VBLK23

Quant Time: Nov 05 06:31:37 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM110420S.M
 Quant Title : VOC Analysis
 QLast Update : Wed Nov 04 16:12:16 2020
 Response via : Initial Calibration

Manual Integrations
 APPROVED

MMDadoda
 11/6/2020 9:37:11 AM



(13) Acetone (T)

4.117min (-0.012) 0.63ug/L

response 642

Ion	Exp%	Act%
43.00	100	100
58.00	34.60	21.18
0.00	0.00	0.00
0.00	0.00	0.00

Quantitation Report (Qedit)




Data Path : Z:\voasrv\HPCHEM1\MSVOA W\Data\VW110420\
 Data File : VW017080.D
 Acq On : 04 Nov 2020 18:04
 Operator : SY/VA
 Sample : VW1104SBL01
 Misc : 5.00G/10ML/MSVOA W/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VBLK23

Quant Time: Nov 05 06:31:37 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM110420S.M
 Quant Title : VOC Analysis
 QLast Update : Wed Nov 04 16:12:16 2020
 Response via : Initial Calibration

Manual Integrations
 APPROVED

MMDadoda
 11/6/2020 9:37:11 AM



(13) Acetone (T)

4.129min (+0.000) 1.53ug/L m

11/10/2020

response 1562

Ion	Exp%	Act%
43.00	100	100
58.00	34.60	8.71
0.00	0.00	0.00
0.00	0.00	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW110420\
Quantitation Report (LSC Reviewed)

Data File : VW017080.D
Acq On : 04 Nov 2020 18:04
Operator : SY/VA
Sample : VW1104SBL01
Misc : 5.00G/10ML/MSVOA_W/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_W
Client Sample ID :
VBLK23

Quant Time: Nov 05 20:45:38 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM110420S.M
Quant Title : VOC Analysis
QLast Update : Wed Nov 04 16:12:16 2020
Response via : Initial Calibration

Manual Integrations
APPROVED

MMDadoda
11/6/2020 9:37:11 AM

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.84	114	542143	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.63	117	495240	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.56	152	241873	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.36	65	179545	26.29	ug/L	0.00
Spiked Amount	25.000	Range 30 - 150	Recovery	=	105.16%	
7) Chloroethane-d5	2.89	69	136401	25.92	ug/L	0.00
Spiked Amount	25.000	Range 30 - 150	Recovery	=	103.68%	
10) 1,1-Dichloroethene-d2	4.02	63	234977	19.24	ug/L	0.00
Spiked Amount	25.000	Range 45 - 110	Recovery	=	76.96%	
20) 2-Butanone-d5	7.07	46	83011	54.36	ug/L	0.00
Spiked Amount	50.000	Range 20 - 135	Recovery	=	108.72%	
24) Chloroform-d	7.65	84	323980	25.60	ug/L	0.00
Spiked Amount	25.000	Range 40 - 150	Recovery	=	102.40%	
26) 1,2-Dichloroethane-d4	8.31	65	170016	27.02	ug/L	0.00
Spiked Amount	25.000	Range 70 - 130	Recovery	=	108.08%	
29) Benzene-d6	8.27	84	678308	26.90	ug/L	0.00
Spiked Amount	25.000	Range 20 - 135	Recovery	=	107.60%	
33) 1,2-Dichloropropane-d6	9.27	67	186020	26.67	ug/L	0.00
Spiked Amount	25.000	Range 70 - 120	Recovery	=	106.68%	
37) Toluene-d8	10.32	98	631018	26.54	ug/L	0.00
Spiked Amount	25.000	Range 30 - 130	Recovery	=	106.16%	
38) trans-1,3-Dichloropropene-	10.58	79	83056	26.41	ug/L	0.00
Spiked Amount	25.000	Range 30 - 135	Recovery	=	105.64%	
39) 2-Hexanone-d5	10.93	63	68902	56.60	ug/L	0.00
Spiked Amount	50.000	Range 20 - 135	Recovery	=	113.20%	
48) 1,1,2,2-Tetrachloroethane-	12.69	84	163266	26.87	ug/L	0.00
Spiked Amount	25.000	Range 45 - 120	Recovery	=	107.48%	
61) 1,2-Dichlorobenzene-d4	13.85	152	235649	27.65	ug/L	0.00
Spiked Amount	25.000	Range 75 - 120	Recovery	=	110.60%	

Target Compounds

13) Acetone	4.13	43	1562m	1.528	ug/L	93
16) Methylene chloride	4.92	84	6911	0.946	ug/L	93
43) 4-Methyl-2-pentanone	10.21	43	5230	1.394	ug/L	95

Qvalue

11/10/20 5/

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW110420\

Quantitation Report (LSC Reviewed)

Data File : VW017080.D

Acq On : 04 Nov 2020 18:04

Operator : SY/VA

Sample : VW1104SBL01

Misc : 5.00G/10ML/MSVOA_W/SOIL

ALS Vial : 14 Sample Multiplier: 1

Instrument :

MSVOA_W

ClientSampleId :

VLBK23

Quant Time: Nov 05 20:45:38 2020

Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM110420S.M

Quant Title : VOC Analysis

QLast Update : Wed Nov 04 16:12:16 2020

Response via : Initial Calibration

Manual Integrations
APPROVED

MMDadoda
11/6/2020 9:37:11 AM

