

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW111020\
 Data File : VW017129.D
 Acq On : 10 Nov 2020 10:13
 Operator : SY/VA
 Sample : VW1110SBL01
 Misc : 5.00G/10ML/MSVOA W/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VBLK27

Quant Time: Nov 11 05:49:47 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM110420S.M
 Quant Title : VOC Analysis
 QLast Update : Wed Nov 11 05:30:56 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.84	114	472107	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.63	117	427481	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.56	152	200380	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.36	65	152982	25.73	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	102.92%
7) Chloroethane-d5	2.89	69	123555	26.96	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	107.84%
10) 1,1-Dichloroethene-d2	4.01	63	197922	18.61	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	74.44%
20) 2-Butanone-d5	7.07	46	64846	48.76	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	97.52%
24) Chloroform-d	7.65	84	293842	26.66	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	106.64%
26) 1,2-Dichloroethane-d4	8.31	65	147339	26.89	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	107.56%
29) Benzene-d6	8.27	84	574918	26.41	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	105.64%
33) 1,2-Dichloropropane-d6	9.27	67	162645	27.01	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	108.04%
37) Toluene-d8	10.32	98	534891	26.06	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	104.24%
38) trans-1,3-Dichloropropene-	10.58	79	68288	25.15	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	100.60%
39) 2-Hexanone-d5	10.93	63	51085	48.62	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	97.24%
48) 1,1,2,2-Tetrachloroethane-	12.69	84	136646	26.05	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	104.20%
61) 1,2-Dichlorobenzene-d4	13.85	152	192432	27.26	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	109.04%

Target Compounds					Ovalue
16) Methylene chloride	4.92	84	6822	1.072 ug/L	96
43) 4-Methyl-2-pentanone	10.21	43	2845	0.879 ug/L	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW111020\
Data File : VW017129.D
Acq On : 10 Nov 2020 10:13
Operator : SY/VA
Sample : VW1110SBL01
Misc : 5.00G/10ML/MSVOA W/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VBLK27

Quant Time: Nov 11 05:49:47 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM110420S.M
Quant Title : VOC Analysis
QLast Update : Wed Nov 11 05:30:56 2020
Response via : Initial Calibration

