

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW111220\
 Data File : VW017184.D
 Acq On : 12 Nov 2020 14:57
 Operator : SY/VA
 Sample : L4693-17
 Misc : 5.02G/10ML/MSVOA W/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BG4E3

Quant Time: Nov 13 04:57:38 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM110420S.M
 Quant Title : VOC Analysis
 QLast Update : Fri Nov 13 04:11:30 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.85	114	371935	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.63	117	326812	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.56	152	128395	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.35	65	113311	24.19	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	96.76%
7) Chloroethane-d5	2.89	69	93775	25.98	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	103.92%
10) 1,1-Dichloroethene-d2	4.02	63	146897	17.53	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	70.12%
20) 2-Butanone-d5	7.09	46	47621	45.46	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	90.92%
24) Chloroform-d	7.65	84	226082	26.04	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	104.16%
26) 1,2-Dichloroethane-d4	8.31	65	115276	26.70	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	106.80%
29) Benzene-d6	8.28	84	445425	26.76	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	107.04%
33) 1,2-Dichloropropane-d6	9.27	67	124377	27.02	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	108.08%
37) Toluene-d8	10.33	98	407875	25.99	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	103.96%
38) trans-1,3-Dichloropropene-	10.58	79	47860	23.06	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	92.24%
39) 2-Hexanone-d5	10.93	63	36760	45.76	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	91.52%
48) 1,1,2,2-Tetrachloroethane-	12.69	84	98376	24.53	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	98.12%
61) 1,2-Dichlorobenzene-d4	13.85	152	115837	25.61	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	102.44%

Target Compounds

					Ovalue
13) Acetone	4.13	43	9118	12.999 ug/L	94
14) Carbon disulfide	4.38	76	24457	1.947 ug/L	99
16) Methylene chloride	4.92	84	8421	1.679 ug/L	100
21) 2-Butanone	7.18	43	2460	2.184 ug/L	87
43) 4-Methyl-2-pentanone	10.21	43	2106	0.851 ug/L #	84
47) Tetrachloroethene	10.87	164	16382	3.777 ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW111220\
Data File : VW017184.D
Acq On : 12 Nov 2020 14:57
Operator : SY/VA
Sample : L4693-17
Misc : 5.02G/10ML/MSVOA W/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BG4E3

Quant Time: Nov 13 04:57:38 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM110420S.M
Quant Title : VOC Analysis
QLast Update : Fri Nov 13 04:11:30 2020
Response via : Initial Calibration

