

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW111220\
 Data File : VW017185.D
 Acq On : 12 Nov 2020 15:21
 Operator : SY/VA
 Sample : VIBLK02
 Misc : 5.00G/10ML/MSVOA W/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VIBLK02

Quant Time: Nov 13 05:00:46 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM110420S.M
 Quant Title : VOC Analysis
 QLast Update : Fri Nov 13 04:11:30 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.84	114	375280	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.63	117	364578	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.56	152	171277	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.36	65	117046	24.76	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	99.04%
7) Chloroethane-d5	2.89	69	97075	26.65	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	106.60%
10) 1,1-Dichloroethene-d2	4.02	63	150619	17.81	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	71.24%
20) 2-Butanone-d5	7.08	46	61011	57.72	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	115.44%
24) Chloroform-d	7.65	84	236340	26.98	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	107.92%
26) 1,2-Dichloroethane-d4	8.31	65	125966	28.92	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	115.68%
29) Benzene-d6	8.27	84	449483	24.21	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	96.84%
33) 1,2-Dichloropropane-d6	9.27	67	130759	25.46	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	101.84%
37) Toluene-d8	10.33	98	413543	23.62	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	94.48%
38) trans-1,3-Dichloropropene-	10.58	79	55046	23.78	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	95.12%
39) 2-Hexanone-d5	10.93	63	48707	54.35	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	108.70%
48) 1,1,2,2-Tetrachloroethane-	12.69	84	122824	27.46	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	109.84%
61) 1,2-Dichlorobenzene-d4	13.85	152	165832	27.48	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	109.92%

Target Compounds					Ovalue
16) Methylene chloride	4.92	84	12937	2.557 ug/L	95
43) 4-Methyl-2-pentanone	10.21	43	2837	1.028 ug/L #	88
47) Tetrachloroethene	10.86	164	1480	0.306 ug/L	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW111220\
Data File : VW017185.D
Acq On : 12 Nov 2020 15:21
Operator : SY/VA
Sample : VIBLK02
Misc : 5.00G/10ML/MSVOA W/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VIBLK02

Quant Time: Nov 13 05:00:46 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM110420S.M
Quant Title : VOC Analysis
QLast Update : Fri Nov 13 04:11:30 2020
Response via : Initial Calibration

