

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW111221\
 Data File : VW020874.D
 Acq On : 12 Nov 2021 23:36
 Operator : SY/VA
 Sample : M4577-03
 Misc : 5.63g/5mL/MSVOA_W/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 SB-3

Quant Time: Nov 15 00:14:13 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W111021S.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 13 07:30:48 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.946	168	71426	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.842	114	109771	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.628	117	91260	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.554	152	36968	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.305	65	38360	64.944	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	129.880%
35) Dibromofluoromethane	7.885	113	35499	57.846	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	115.700%
50) Toluene-d8	10.323	98	117652	49.001	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	98.000%
62) 4-Bromofluorobenzene	12.615	95	38252	44.158	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	88.320%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW111221\
Data File : VW020874.D
Acq On : 12 Nov 2021 23:36
Operator : SY/VA
Sample : M4577-03
Misc : 5.63g/5mL/MSVOA_W/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
SB-3

Quant Time: Nov 15 00:14:13 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W111021S.M
Quant Title : SW846 8260
QLast Update : Sat Nov 13 07:30:48 2021
Response via : Initial Calibration

