

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW111620\
 Data File : VW017257.D
 Acq On : 16 Nov 2020 15:20
 Operator : SY/VA
 Sample : L4739-13
 Misc : 6.74G/10ML/MSVOA W/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 GB100

Manual Integrations
 APPROVED

MMDadoda
 11/17/2020 6:53:39 PM

Quant Time: Nov 17 06:59:05 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM110420S.M
 Quant Title : VOC Analysis
 QLast Update : Tue Nov 17 06:08:23 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.84	114	394802	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.63	117	347494	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.56	152	139848	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.36	65	92880	18.68	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	74.72%
7) Chloroethane-d5	2.89	69	80707	21.06	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	84.24%
10) 1,1-Dichloroethene-d2	4.02	63	140227	15.77	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	63.08%
20) 2-Butanone-d5	7.08	46	58522	52.63	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	105.26%
24) Chloroform-d	7.65	84	222687	24.16	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	96.64%
26) 1,2-Dichloroethane-d4	8.31	65	125754	27.44	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	109.76%
29) Benzene-d6	8.27	84	418230	23.63	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	94.52%
33) 1,2-Dichloropropane-d6	9.27	67	120025	24.52	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	98.08%
37) Toluene-d8	10.32	98	394212	23.63	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	94.52%
38) trans-1,3-Dichloropropene-	10.58	79	53360	24.18	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	96.72%
39) 2-Hexanone-d5	10.93	63	47839	56.01	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	112.02%
48) 1,1,2,2-Tetrachloroethane-	12.69	84	112289	26.34	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	105.36%
61) 1,2-Dichlorobenzene-d4	13.85	152	118794	24.11	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	96.44%

Target Compounds

					Ovalue
5) Vinyl chloride	2.37	62	2905	0.551 ug/L #	1
13) Acetone	4.13	43	4931	6.622 ug/L	65
16) Methylene chloride	4.93	84	4525m	0.850 ug/L	
22) cis-1,2-Dichloroethene	7.17	96	67446	12.576 ug/L	97
35) Trichloroethene	9.09	95	33810	6.615 ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW111620\
Data File : VW017257.D
Acq On : 16 Nov 2020 15:20
Operator : SY/VA
Sample : L4739-13
Misc : 6.74G/10ML/MSVOA W/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
GB100

Manual Integrations
APPROVED

MMDadoda
11/17/2020 6:53:39 PM

Quant Time: Nov 17 06:59:05 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM110420S.M
Quant Title : VOC Analysis
QLast Update : Tue Nov 17 06:08:23 2020
Response via : Initial Calibration

