

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW111720\
 Data File : VW017315.D
 Acq On : 18 Nov 2020 03:27
 Operator : SY/VA
 Sample : L4738-21
 Misc : 2.66G/10ML/MSVOA W/SOIL
 ALS Vial : 52 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :

Quant Time: Nov 18 06:06:19 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM110420S.M
 Quant Title : VOC Analysis
 QLast Update : Wed Nov 18 04:50:20 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.85	114	225624	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.63	117	94583	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.56	152	13550	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.35	65	90880	31.98	ug/L	-0.01
Spiked Amount	25.000	Range	30 - 150	Recovery	=	127.92%
7) Chloroethane-d5	2.89	69	81883	37.39	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	149.56%
10) 1,1-Dichloroethene-d2	4.01	63	112657	22.16	ug/L	-0.01
Spiked Amount	25.000	Range	45 - 110	Recovery	=	88.64%
20) 2-Butanone-d5	7.09	46	41275	64.95	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	129.90%
24) Chloroform-d	7.65	84	169879	32.25	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	129.00%
26) 1,2-Dichloroethane-d4	8.31	65	75896	28.98	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	115.92%
29) Benzene-d6	8.27	84	292180	60.66	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	242.64%#
33) 1,2-Dichloropropane-d6	9.27	67	80047	60.08	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	240.32%#
37) Toluene-d8	10.33	98	180514	39.75	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	159.00%#
38) trans-1,3-Dichloropropene-	10.58	79	17349	28.88	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	115.52%
39) 2-Hexanone-d5	10.93	63	16981	73.04	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	146.08%#
48) 1,1,2,2-Tetrachloroethane-	12.69	84	27796	23.95	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	95.80%
61) 1,2-Dichlorobenzene-d4	13.85	152	10732	22.48	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	89.92%

Target Compounds					Ovalue
13) Acetone	4.14	43	6088	14.307 ug/L	64
16) Methylene chloride	4.91	84	12941m	4.254 ug/L	
44) Toluene	10.39	91	1960	0.359 ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW11720\
Data File : VW017315.D
Acq On : 18 Nov 2020 03:27
Operator : SY/VA
Sample : L4738-21
Misc : 2.66G/10ML/MSVOA W/SOIL
ALS Vial : 52 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :

Quant Time: Nov 18 06:06:19 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM110420S.M
Quant Title : VOC Analysis
QLast Update : Wed Nov 18 04:50:20 2020
Response via : Initial Calibration

