

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW111720\
 Data File : VW017312.D
 Acq On : 18 Nov 2020 02:16
 Operator : SY/VA
 Sample : L4738-18
 Misc : 5.67G/10ML/MSVOA W/SOIL
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 GB0Y5

Quant Time: Nov 18 05:57:52 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM110420S.M
 Quant Title : VOC Analysis
 QLast Update : Wed Nov 18 04:50:20 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.84	114	397600	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.63	117	333795	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.56	152	99788	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.35	65	89133	17.80	ug/L	-0.01
Spiked Amount	25.000	Range	30 - 150	Recovery	=	71.20%
7) Chloroethane-d5	2.89	69	80387	20.83	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	83.32%
10) 1,1-Dichloroethene-d2	4.02	63	130136	14.53	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	58.12%
20) 2-Butanone-d5	7.08	46	61620	55.02	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	110.04%
24) Chloroform-d	7.65	84	227985	24.56	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	98.24%
26) 1,2-Dichloroethane-d4	8.31	65	128228	27.79	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	111.16%
29) Benzene-d6	8.27	84	429669	25.28	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	101.12%
33) 1,2-Dichloropropane-d6	9.27	67	128358	27.30	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	109.20%
37) Toluene-d8	10.33	98	373314	23.29	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	93.16%
38) trans-1,3-Dichloropropene-	10.58	79	52834	24.92	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	99.68%
39) 2-Hexanone-d5	10.93	63	53558	65.27	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	130.54%
48) 1,1,2,2-Tetrachloroethane-	12.69	84	120149	29.34	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	117.36%
61) 1,2-Dichlorobenzene-d4	13.85	152	94288	26.82	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	107.28%

Target Compounds

					Ovalue
13) Acetone	4.13	43	6537	8.718 ug/L	100
16) Methylene chloride	4.93	84	12312	2.297 ug/L	90
34) Benzene	8.33	78	7727	0.441 ug/L	100
36) Methylcyclohexane	9.34	83	2225	0.282 ug/L	90
44) Toluene	10.39	91	14909	0.773 ug/L	97
54) m,p-Xylene	11.84	106	2369	0.281 ug/L	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW11720\
Data File : VW017312.D
Acq On : 18 Nov 2020 02:16
Operator : SY/VA
Sample : L4738-18
Misc : 5.67G/10ML/MSVOA W/SOIL
ALS Vial : 49 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
GB0Y5

Quant Time: Nov 18 05:57:52 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM110420S.M
Quant Title : VOC Analysis
QLast Update : Wed Nov 18 04:50:20 2020
Response via : Initial Calibration

