

Data File : VW017369.D
Acq On : 19 Nov 2020 14:12
Operator : SY/VA
Sample : L4793-03
Misc : 6.06G/10ML/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :

Quant Time: Nov 20 03:39:07 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM111820S.M
Quant Title : VOC Analysis
QLast Update : Fri Nov 20 01:18:11 2020
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.85	114	398095	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.63	117	361260	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.56	152	160200	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.36	65	129803	22.02	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	88.08%
7) Chloroethane-d5	2.89	69	102260	21.84	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	87.36%
10) 1,1-Dichloroethene-d2	4.02	63	164595	15.96	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	63.84%
20) 2-Butanone-d5	7.08	46	33315	34.42	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	68.84%
24) Chloroform-d	7.65	84	219169	20.96	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	83.84%
26) 1,2-Dichloroethane-d4	8.31	65	101974	20.18	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	80.72%
29) Benzene-d6	8.27	84	433804	21.30	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	85.20%
33) 1,2-Dichloropropane-d6	9.27	67	116572	21.49	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	85.96%
37) Toluene-d8	10.32	98	405241	21.14	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	84.56%
38) trans-1,3-Dichloropropene-	10.58	79	45212	19.14	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	76.56%
39) 2-Hexanone-d5	10.93	63	25426	34.95	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	69.90%
48) 1,1,2,2-Tetrachloroethane-	12.69	84	78867	19.18	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	76.72%
61) 1,2-Dichlorobenzene-d4	13.85	152	130030	21.99	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	87.96%

Target Compounds					Qvalue
13) Acetone	4.14	43	1589m	2.250 ug/L	
16) Methylene chloride	4.92	84	17716	2.602 ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VW017369.D
Acq On : 19 Nov 2020 14:12
Operator : SY/VA
Sample : L4793-03
Misc : 6.06G/10ML/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :

Quant Time: Nov 20 03:39:07 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM111820S.M
Quant Title : VOC Analysis
QLast Update : Fri Nov 20 01:18:11 2020
Response via : Initial Calibration

