


Data Path : Z:\voasrv\HPCHEM1\MSVOA W\Data\VW111920\
 Data File : VW017356.D
 Acq On : 19 Nov 2020 08:24
 Operator : SY/VA
 Sample : BFB10
 Misc : 5.00G/10ML/MSVOA W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BFB10

Integration File: LSCINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM111820S.M
 Title : VOC Analysis
 Last Update : Fri Nov 20 09:29:51 2020



AutoFind: Scans 1757, 1758, 1759; Background Corrected with Scan 1749

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.8	27101	PASS
75	95	30	80	50.4	76744	PASS
95	95	100	100	100.0	152256	PASS
96	95	5	9	6.4	9799	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	120	92.6	141045	PASS
175	174	5	9	7.5	10556	PASS
176	174	95	101	95.7	134946	PASS
177	176	5	9	7.2	9747	PASS

m/z	Abundance
36.00	1379.0
37.00	7666.0
38.00	6486.0
39.10	2370.0
40.10	324.0
43.00	193.0
44.00	1340.0
45.00	1629.0
46.20	122.0
47.00	1498.0
48.00	890.0
49.00	6587.0
50.00	28800.0
51.00	8477.0
52.10	432.0
52.90	67.0
55.00	508.0
56.00	2405.0
57.00	4594.0
58.10	208.0
59.00	378.0
60.00	1311.0
61.00	8119.0
62.00	7499.0
63.00	5481.0
64.00	543.0
65.00	655.0
66.90	419.0
68.00	15495.0
69.00	15638.0
70.00	1186.0
71.00	93.0
72.00	813.0
73.00	8858.0
74.00	27256.0
75.00	76552.0
76.00	6678.0
77.10	938.0
77.90	543.0
78.90	5628.0
79.90	1697.0
80.90	5797.0
81.90	1219.0
82.90	196.0
85.80	149.0
87.00	6638.0
88.00	6276.0
91.00	706.0
92.00	4451.0
93.00	7271.0
94.00	19416.0
95.00	147968.0
96.00	10311.0
97.00	409.0
103.00	580.0
103.90	871.0
104.80	443.0
105.80	740.0
107.00	113.0
109.90	268.0
110.80	341.0
112.00	233.0
112.90	231.0
115.00	391.0
115.90	680.0
116.90	955.0
117.90	823.0
118.90	1405.0
120.00	68.0
120.80	110.0
123.90	78.0
124.90	740.0
125.80	292.0
126.50	89.0
127.90	568.0
128.90	273.0
129.90	594.0
131.00	323.0
132.10	124.0
133.00	2621.0
134.90	471.0
136.10	90.0
137.10	265.0
139.00	96.0
140.90	1815.0
142.00	214.0
142.90	1771.0
144.00	102.0
144.90	399.0
145.90	322.0
146.90	523.0
147.90	428.0

Instrument :
MSVOA_W
ClientSampleId :
BFB10

149.00	297.0
150.00	198.0
150.90	115.0
151.90	62.0
152.90	147.0
153.90	96.0
154.90	384.0
156.00	130.0
157.00	282.0
157.90	72.0
158.80	204.0
161.00	345.0
162.90	347.0
164.90	379.0
166.00	55.0
166.80	52.0
170.50	169.0
171.80	1000.0
172.20	946.0
174.00	122080.0
175.00	9817.0
176.00	117544.0
177.00	8720.0
177.90	427.0
178.90	477.0
179.80	82.0
180.80	59.0
189.10	193.0
189.90	89.0
191.00	1979.0
192.10	425.0
193.00	3105.0
194.10	599.0
195.00	413.0
203.10	271.0
205.00	612.0
205.90	149.0
207.10	1747.0
208.00	454.0
208.80	298.0
218.70	84.0
220.90	128.0
235.00	267.0
249.00	1790.0
249.90	541.0
251.00	782.0
251.90	363.0
265.10	2142.0
266.10	436.0
267.00	489.0
268.10	231.0
269.20	88.0
278.20	88.0
279.10	241.0
281.10	21592.0
282.10	7448.0
283.00	4039.0
284.00	819.0
284.90	216.0

Instrument :
MSVOA_W
ClientSampleId :
BFB10

m/z	Abundance
36.00	1332.0
37.10	7089.0
38.00	7155.0
39.10	2416.0
40.00	268.0
40.90	65.0
43.10	145.0
44.00	1282.0
45.00	1495.0
46.00	200.0
47.10	1667.0
48.00	1010.0
49.00	7179.0
50.00	30024.0
51.00	9140.0
51.90	414.0
55.00	695.0
56.00	2766.0
57.00	4839.0
57.90	210.0
59.00	230.0
60.00	1630.0
61.00	8584.0
62.00	8102.0
63.00	6257.0
64.10	626.0
65.00	650.0
67.00	448.0
68.00	16808.0
69.00	17880.0
70.10	1385.0
71.10	76.0
72.10	995.0
73.00	9249.0
74.10	31240.0
75.00	87448.0
76.10	7767.0
77.10	828.0
78.00	769.0
79.00	6246.0
80.00	1821.0
80.90	6767.0
81.90	1395.0
83.00	182.0
85.80	259.0
87.00	7585.0
88.00	6434.0
90.90	800.0
92.00	5531.0
93.00	7794.0
94.00	21240.0
95.00	175104.0
96.00	10414.0
97.00	385.0
102.80	264.0
103.20	254.0
103.90	999.0
104.90	361.0
105.90	902.0
106.90	202.0
110.00	195.0
111.00	251.0
112.10	218.0
113.00	196.0
115.00	285.0
115.90	533.0
116.90	1350.0
118.00	935.0
119.00	1241.0
120.00	133.0
123.90	85.0
125.20	449.0
125.90	281.0
127.90	689.0
128.80	442.0
129.80	665.0
130.90	401.0
131.90	62.0
132.90	1549.0
134.00	271.0
134.90	490.0
136.00	70.0
136.90	397.0
138.70	91.0
140.00	139.0
141.00	1936.0
143.00	2103.0
143.80	105.0
144.20	63.0
145.10	338.0
145.90	300.0
146.90	332.0

Instrument :
MSVOA_W
ClientSampleId :
BFB10

147.90	411.0
149.00	178.0
150.00	296.0
152.00	63.0
152.90	156.0
153.90	154.0
154.90	424.0
155.90	169.0
157.00	354.0
157.90	128.0
158.80	193.0
161.00	302.0
162.90	230.0
165.00	237.0
166.10	56.0
168.90	53.0
170.30	109.0
171.10	189.0
172.00	1166.0
173.90	160256.0
175.00	11697.0
176.00	150720.0
176.90	10810.0
177.90	376.0
179.10	349.0
179.80	64.0
180.80	56.0
188.90	229.0
189.90	87.0
190.90	1515.0
192.10	293.0
193.00	2002.0
194.00	312.0
195.00	199.0
203.00	293.0
204.20	72.0
205.00	324.0
205.90	106.0
207.10	1233.0
208.10	428.0
208.90	105.0
220.90	95.0
235.00	163.0
249.00	1715.0
250.00	246.0
251.00	553.0
252.00	163.0
265.00	1524.0
265.90	304.0
267.10	342.0
268.10	115.0
278.30	56.0
279.00	133.0
281.10	19000.0
282.00	4963.0
283.00	3305.0
283.90	423.0
285.00	166.0

Instrument :
MSVOA_W
ClientSampleId :
BFB10

m/z	Abundance
36.00	1105.0
37.10	5936.0
38.00	5215.0
39.00	2062.0
40.10	208.0
40.90	80.0
42.90	173.0
44.10	902.0
45.00	1242.0
45.90	121.0
46.90	1483.0
48.00	975.0
49.10	4924.0
50.00	22480.0
51.00	7285.0
52.00	365.0
54.80	298.0
56.00	2055.0
57.00	3922.0
58.20	136.0
59.00	108.0
60.10	1427.0
61.00	6645.0
62.00	6617.0
63.00	4907.0
64.00	545.0
64.90	336.0
67.10	539.0
68.00	12907.0
69.00	13286.0
70.10	1099.0
71.10	124.0
72.00	813.0
73.00	6268.0
74.00	23632.0
75.00	66232.0
76.00	5720.0
76.80	893.0
77.90	665.0
78.90	4948.0
80.00	1398.0
80.90	5149.0
81.90	1189.0
83.10	141.0
86.00	139.0
87.00	6039.0
88.00	5587.0
91.10	591.0
92.00	4332.0
93.00	5899.0
94.00	17608.0
95.00	133696.0
96.10	8673.0
97.00	285.0
103.10	96.0
103.90	641.0
104.80	229.0
106.00	776.0
106.70	190.0
109.80	108.0
111.00	264.0
111.80	170.0
112.70	125.0
114.80	237.0
116.00	672.0
116.90	995.0
117.90	681.0
119.00	880.0
123.90	90.0
124.80	107.0
125.30	139.0
125.90	140.0
127.90	494.0
129.00	269.0
129.90	577.0
130.90	182.0
131.90	81.0
133.00	719.0
134.00	215.0
134.90	498.0
136.80	310.0
138.70	88.0
140.00	171.0
141.00	1757.0
142.00	301.0
143.00	1841.0
143.90	126.0
145.00	269.0
145.70	225.0
146.90	204.0
148.00	325.0
148.90	131.0

Instrument :
MSVOA_W
ClientSampleId :
BFB10

149.90	233.0
151.80	77.0
152.90	132.0
154.00	96.0
155.00	387.0
156.00	155.0
157.00	236.0
157.90	102.0
159.00	240.0
160.90	236.0
162.90	92.0
165.10	69.0
169.50	57.0
170.80	146.0
171.80	852.0
173.90	140800.0
175.00	10154.0
175.90	136576.0
177.00	9711.0
178.00	219.0
179.00	197.0
189.00	77.0
190.90	542.0
191.80	104.0
193.00	912.0
193.90	176.0
194.90	135.0
203.30	76.0
204.90	130.0
207.00	634.0
207.90	122.0
209.00	85.0
235.00	90.0
249.00	602.0
249.90	114.0
250.90	312.0
252.10	93.0
265.00	687.0
266.10	172.0
266.90	94.0
267.20	101.0
268.10	56.0
279.10	108.0
279.80	103.0
281.10	10289.0
282.00	2334.0
283.10	1439.0
283.90	254.0
285.10	158.0

Instrument :
MSVOA_W
ClientSampleId :
BFB10

Scan 1749 (12.567 min): VW017356.D
BFB10

m/z	Abundance
39.90	109.0
41.00	70.0
43.90	75.0
44.20	76.0
281.30	57.0

Instrument :
MSVOA_W
ClientSampleId :
BFB10