

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW122021\
 Data File : VW021506.D
 Acq On : 20 Dec 2021 18:24
 Operator : SY/VA
 Sample : VIBLK
 Misc : 5.00g/10.0mL/MSVOA_W/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VIBLK555

Quant Time: Dec 20 21:31:52 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM120921SMA.M
 Quant Title : SFAM01.0
 QLast Update : Sat Dec 18 00:55:20 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	172588	25.000	ug/L	# 0.00
28) Chlorobenzene-d5	11.634	117	168793	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.560	152	94992	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	45829	14.580	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	58.320%
7) Chloroethane-d5	2.886	69	30133	14.501	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	58.000%
11) 1,1-Dichloroethene-d2	4.013	63	70451	16.812	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	67.240%
21) 2-Butanone-d5	7.092	46	22639	51.095	ug/L	0.02
Spiked Amount	50.000	Range	20 - 135	Recovery	=	102.180%
24) Chloroform-d	7.647	84	107873	22.428	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	89.720%
26) 1,2-Dichloroethane-d4	8.305	65	57338	22.412	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	89.640%
32) Benzene-d6	8.275	84	225186	24.339	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	97.360%
36) 1,2-Dichloropropane-d6	9.275	67	62059	25.011	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	100.040%
41) Toluene-d8	10.323	98	222436	23.530	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	94.120%
43) trans-1,3-Dichloroprop...	10.579	79	26168	22.904	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	91.600%
47) 2-Hexanone-d5	10.927	63	19997	54.500	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	109.000%
56) 1,1,2,2-Tetrachloroeth...	12.695	84	47837	22.427	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	89.720%
66) 1,2-Dichlorobenzene-d4	13.853	152	88891	24.146	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	96.600%
Target Compounds					Qvalue	
53) m,p-Xylene	11.835	106	4694	0.947	ug/L	78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW122021\
Data File : VW021506.D
Acq On : 20 Dec 2021 18:24
Operator : SY/VA
Sample : VIBLK
Misc : 5.00g/10.0mL/MSVOA_W/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VIBLK555

Quant Time: Dec 20 21:31:52 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM120921SMA.M
Quant Title : SFAM01.0
QLast Update : Sat Dec 18 00:55:20 2021
Response via : Initial Calibration

