

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW061722\
 Data File : VW023494.D
 Acq On : 17 Jun 2022 15:48
 Operator : SY/VA
 Sample : VIBLK
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :

Quant Time: Jun 17 23:31:43 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM061022SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Jun 17 23:30:09 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	332665	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.628	117	316274	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	144575	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	120221	22.747	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	91.000%
7) Chloroethane-d5	2.879	69	78367	20.172	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	80.680%
11) 1,1-Dichloroethene-d2	4.013	63	106604	14.451	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	57.800%
21) 2-Butanone-d5	7.080	46	58466	54.182	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	108.360%
24) Chloroform-d	7.647	84	201179	21.969	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	87.880%
26) 1,2-Dichloroethane-d4	8.305	65	113108	22.934	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	91.720%
32) Benzene-d6	8.275	84	367797	21.932	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	87.720%
36) 1,2-Dichloropropane-d6	9.274	67	114666	23.209	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	92.840%
41) Toluene-d8	10.323	98	341639	20.857	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	83.440%
43) trans-1,3-Dichloroprop...	10.579	79	47719	21.161	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	84.640%
47) 2-Hexanone-d5	10.920	63	45686	57.869	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	115.740%
56) 1,1,2,2-Tetrachloroeth...	12.688	84	112476	24.375	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	97.480%
66) 1,2-Dichlorobenzene-d4	13.847	152	129575	24.827	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	99.320%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW061722\
Data File : VW023494.D
Acq On : 17 Jun 2022 15:48
Operator : SY/VA
Sample : VIBLK
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :

Quant Time: Jun 17 23:31:43 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM061022SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Jun 17 23:30:09 2022
Response via : Initial Calibration

