

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW061722\
 Data File : VW023492.D
 Acq On : 17 Jun 2022 14:56
 Operator : SY/VA
 Sample : N3345-05
 Misc : 5.02g/10mL/MSVOA_W/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VHBLK002

Quant Time: Jun 17 23:31:14 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM061022SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Jun 17 23:30:09 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	327481	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.628	117	311445	25.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	138730	25.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	116325	22.36	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	89.44%
7) Chloroethane-d5	2.879	69	78739	20.59	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	82.36%
11) 1,1-Dichloroethene-d2	4.013	63	110079	15.16	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	60.64%
21) 2-Butanone-d5	7.080	46	53849	50.69	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	101.38%
24) Chloroform-d	7.647	84	212449	23.57	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	94.28%
26) 1,2-Dichloroethane-d4	8.305	65	114594	23.60	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	94.40%
32) Benzene-d6	8.275	84	385544	23.35	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	93.40%
36) 1,2-Dichloropropane-d6	9.275	67	122253	25.13	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	100.52%
41) Toluene-d8	10.323	98	363268	22.52	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	90.08%
43) trans-1,3-Dichloroprop...	10.579	79	48351	21.77	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	87.08%
47) 2-Hexanone-d5	10.927	63	41516	53.40	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	106.80%
56) 1,1,2,2-Tetrachloroeth...	12.689	84	109643	24.13	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	96.52%
66) 1,2-Dichlorobenzene-d4	13.853	152	128067	25.57	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	102.28%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW061722\
Data File : VW023492.D
Acq On : 17 Jun 2022 14:56
Operator : SY/VA
Sample : N3345-05
Misc : 5.02g/10mL/MSVOA_W/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VHBLK002

Quant Time: Jun 17 23:31:14 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM061022SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Jun 17 23:30:09 2022
Response via : Initial Calibration

