

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW061722\
 Data File : VW023491.D
 Acq On : 17 Jun 2022 14:32
 Operator : SY/VA
 Sample : N3292-04
 Misc : 5.08g/10mL/MSVOA_W/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VHBLK001

Quant Time: Jun 17 23:30:58 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM061022SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Jun 17 23:30:09 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	339291	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.628	117	322493	25.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	144610	25.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	111508	20.69	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	82.76%
7) Chloroethane-d5	2.885	69	78590	19.83	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	79.32%
11) 1,1-Dichloroethene-d2	4.019	63	106609	14.17	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	56.68%
21) 2-Butanone-d5	7.080	46	55655	50.57	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	101.14%
24) Chloroform-d	7.647	84	205564	22.01	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	88.04%
26) 1,2-Dichloroethane-d4	8.305	65	111927	22.25	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	89.00%
32) Benzene-d6	8.275	84	369560	21.61	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	86.44%
36) 1,2-Dichloropropane-d6	9.275	67	115676	22.96	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	91.84%
41) Toluene-d8	10.323	98	341534	20.45	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	81.80%
43) trans-1,3-Dichloroprop...	10.579	79	47643	20.72	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	82.88%
47) 2-Hexanone-d5	10.927	63	42045	52.23	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	104.46%
56) 1,1,2,2-Tetrachloroeth...	12.689	84	108028	22.96	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	91.84%
66) 1,2-Dichlorobenzene-d4	13.853	152	128907	24.69	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	98.76%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW061722\
Data File : VW023491.D
Acq On : 17 Jun 2022 14:32
Operator : SY/VA
Sample : N3292-04
Misc : 5.08g/10mL/MSVOA_W/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VHBLK001

Quant Time: Jun 17 23:30:58 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM061022SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Jun 17 23:30:09 2022
Response via : Initial Calibration

