

Data File : VW006386.D
Acq On : 26 Oct 2018 03:52
Operator : SY/AP
Sample : J5610-07RE
Misc : 5.08G/5ML/MSVOA_W/SOIL
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
SB-4(0-2)RE

Quant Time: Oct 26 08:09:29 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W101518S.M
Quant Title : SW846 8260
QLast Update : Thu Oct 18 17:01:24 2018
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	37602	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	55698	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	42384	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	14540	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.30	65	22076	70.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	141.64%	
35) Dibromofluoromethane	7.88	113	18687	56.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	113.56%	
50) Toluene-d8	10.33	98	57660	46.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.10%	
62) 4-Bromofluorobenzene	12.62	95	15716	33.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	66.84%	
Target Compounds						
20) Methylene Chloride	4.92	84	2755	6.951	ug/l #	Qvalue 75

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VW006386.D
Acq On : 26 Oct 2018 03:52
Operator : SY/AP
Sample : J5610-07RE
Misc : 5.08G/5ML/MSVOA_W/SOIL
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
SB-4(0-2)RE

Quant Time: Oct 26 08:09:29 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W101518S.M
Quant Title : SW846 8260
QLast Update : Thu Oct 18 17:01:24 2018
Response via : Initial Calibration

