

Data File : VW006376.D
 Acq On : 25 Oct 2018 23:07
 Operator : SY/AP
 Sample : VSTDCCC050
 Misc : 5.00G/5ML/MSVOA_W/SOIL
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_W
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 26 07:30:38 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W101518S.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 18 17:01:24 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	79	0.00
2 T	Dichlorodifluoromethane	0.247	0.259	-4.9	80	0.00
3 P	Chloromethane	0.318	0.339	-6.6	89	0.00
4 C	Vinyl Chloride	0.455	0.482	-5.9#	84	0.00
5 T	Bromomethane	0.414	0.436	-5.3	88	0.00
6 T	Chloroethane	0.326	0.364	-11.7	89	0.00
7 T	Trichlorofluoromethane	0.424	0.465	-9.7	86	0.00
8 T	Diethyl Ether	0.225	0.247	-9.8	83	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.444	0.487	-9.7	90	0.00
10 T	Methyl Iodide	0.739	0.841	-13.8	92	0.00
11 T	Tert butyl alcohol	0.028	0.027	3.6	86	0.00
12 CM	1,1-Dichloroethene	0.418	0.457	-9.3#	91	0.00
13 T	Acrolein	0.027	0.024	11.1	78	0.00
14 T	Allyl chloride	0.597	0.635	-6.4	87	0.00
15 T	Acrylonitrile	0.081	0.095	-17.3	97	0.00
16 T	Acetone	0.089	0.089	0.0	100	0.00
17 T	Carbon Disulfide	1.261	1.255	0.5	80	0.00
18 T	Methyl Acetate	0.209	0.250	-19.6	98	0.00
19 T	Methyl tert-butyl Ether	0.636	0.721	-13.4	93	0.00
20 T	Methylene Chloride	0.484	0.501	-3.5	94	0.00
21 T	trans-1,2-Dichloroethene	0.466	0.528	-13.3	93	0.00
22 T	Diisopropyl ether	1.180	1.339	-13.5	93	0.00
23 T	Vinyl Acetate	0.697	0.776	-11.3	92	0.00
24 P	1,1-Dichloroethane	0.762	0.886	-16.3	94	0.00
25 T	2-Butanone	0.112	0.128	-14.3	98	0.00
26 T	2,2-Dichloropropane	0.537	0.544	-1.3	86	0.00
27 T	cis-1,2-Dichloroethene	0.515	0.605	-17.5	96	0.00
28 T	Bromochloromethane	0.294	0.354	-20.4#	92	0.00
29 T	Tetrahydrofuran	0.069	0.077	-11.6	94	0.00
30 C	Chloroform	0.830	0.978	-17.8#	97	0.00
31 T	Cyclohexane	0.759	0.751	1.1	86	0.00
32 T	1,1,1-Trichloroethane	0.749	0.827	-10.4	91	0.00
33 S	1,2-Dichloroethane-d4	0.415	0.419	-1.0	81	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	80	0.00
35 S	Dibromofluoromethane	0.295	0.301	-2.0	82	0.00
36 T	1,1-Dichloropropene	0.461	0.509	-10.4	92	0.00
37 T	Ethyl Acetate	0.171	0.194	-13.5	99	0.00
38 T	Carbon Tetrachloride	0.505	0.539	-6.7	88	0.00
39 T	Methylcyclohexane	0.613	0.653	-6.5	89	0.00
40 TM	Benzene	1.264	1.452	-14.9	96	0.00
41 T	Methacrylonitrile	0.103	0.111	-7.8	88	0.00
42 TM	1,2-Dichloroethane	0.371	0.433	-16.7	97	0.00
43 T	Isopropyl Acetate	0.350	0.385	-10.0	93	0.00
44 TM	Trichloroethene	0.388	0.443	-14.2	95	0.00
45 C	1,2-Dichloropropane	0.303	0.346	-14.2#	95	0.00

Data File : VW006376.D
 Acq On : 25 Oct 2018 23:07
 Operator : SY/AP
 Sample : VSTDCCC050
 Misc : 5.00G/5ML/MSVOA_W/SOIL
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_W
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 26 07:30:38 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W101518S.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 18 17:01:24 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.172	0.202	-17.4	97	0.00
47 T	Bromodichloromethane	0.418	0.476	-13.9	94	0.00
48 T	Methyl methacrylate	0.166	0.186	-12.0	92	0.00
49 T	1,4-Dioxane	0.003	0.003	0.0	94	0.00
50 S	Toluene-d8	1.124	1.099	2.2	79	0.00
51 T	4-Methyl-2-Pentanone	0.175	0.199	-13.7	96	0.00
52 CM	Toluene	0.848	0.969	-14.3#	96	0.00
53 T	t-1,3-Dichloropropene	0.416	0.469	-12.7	92	0.00
54 T	cis-1,3-Dichloropropene	0.478	0.536	-12.1	92	0.00
55 T	1,1,2-Trichloroethane	0.241	0.285	-18.3	99	0.00
56 T	Ethyl methacrylate	0.296	0.344	-16.2	97	0.00
57 T	1,3-Dichloropropane	0.399	0.468	-17.3	97	0.00
58 T	2-Chloroethyl Vinyl ether	0.113	0.121	-7.1	92	0.00
59 T	2-Hexanone	0.125	0.137	-9.6	94	0.00
60 T	Dibromochloromethane	0.301	0.346	-15.0	93	0.00
61 T	1,2-Dibromoethane	0.241	0.287	-19.1	98	0.00
62 S	4-Bromofluorobenzene	0.422	0.428	-1.4	81	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	80	0.00
64 T	Tetrachloroethene	0.388	0.441	-13.7	95	0.00
65 PM	Chlorobenzene	1.070	1.244	-16.3	97	0.00
66 T	1,1,1,2-Tetrachloroethane	0.380	0.442	-16.3	95	0.00
67 C	Ethyl Benzene	1.855	2.116	-14.1#	95	0.00
68 T	m/p-Xylenes	0.733	0.838	-14.3	94	0.00
69 T	o-Xylene	0.699	0.806	-15.3	95	0.00
70 T	Styrene	1.136	1.325	-16.6	95	0.00
71 P	Bromoform	0.212	0.237	-11.8	89	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	80	0.00
73 T	Isopropylbenzene	3.532	3.932	-11.3	94	0.00
74 T	N-amyl acetate	0.686	0.723	-5.4	89	0.00
75 P	1,1,2,2-Tetrachloroethane	0.537	0.635	-18.2	98	0.00
76 T	1,2,3-Trichloropropane	0.382	0.386	-1.0	88	0.00
77 T	Bromobenzene	0.865	0.997	-15.3	97	0.00
78 T	n-propylbenzene	4.110	4.567	-11.1	93	0.00
79 T	2-Chlorotoluene	2.309	2.587	-12.0	93	0.00
80 T	1,3,5-Trimethylbenzene	2.978	3.313	-11.2	93	0.00
81 T	trans-1,4-Dichloro-2-butene	0.160	0.165	-3.1	85	0.00
82 T	4-Chlorotoluene	2.419	2.731	-12.9	94	0.00
83 T	tert-Butylbenzene	2.664	2.989	-12.2	94	0.00
84 T	1,2,4-Trimethylbenzene	3.013	3.366	-11.7	93	0.00
85 T	sec-Butylbenzene	3.764	4.176	-10.9	92	0.00
86 T	p-Isopropyltoluene	3.404	3.811	-12.0	92	0.00
87 T	1,3-Dichlorobenzene	1.722	1.968	-14.3	96	0.00
88 T	1,4-Dichlorobenzene	1.714	1.954	-14.0	96	0.00
89 T	n-Butylbenzene	3.114	3.415	-9.7	92	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_W\DATA\VW102718\
Evaluate Continuing Calibration Report

Data File : VW006376.D
Acq On : 25 Oct 2018 23:07
Operator : SY/AP
Sample : VSTDCCC050
Misc : 5.00G/5ML/MSVOA_W/SOIL
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_W
LabSampleId :
VSTDCCC050

Quant Time: Oct 26 07:30:38 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W101518S.M
Quant Title : SW846 8260
QLast Update : Thu Oct 18 17:01:24 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.577	0.615	-6.6	85	0.00
91 T	1,2-Dichlorobenzene	1.536	1.759	-14.5	96	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.096	0.105	-9.4	91	0.00
93 T	1,2,4-Trichlorobenzene	1.182	1.338	-13.2	98	0.00
94 T	Hexachlorobutadiene	0.742	0.818	-10.2	94	0.00
95 T	Naphthalene	1.966	2.283	-16.1	99	0.00
96 T	1,2,3-Trichlorobenzene	1.022	1.173	-14.8	98	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6