

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW010722\
 Data File : VW021770.D
 Acq On : 07 Jan 2022 10:42
 Operator : SY/VA
 Sample : VW0107SBL01
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VBLK437

Quant Time: Jan 07 22:01:23 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM122821SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Jan 07 21:59:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.848	114	149745	25.000	ug/L	# 0.00
28) Chlorobenzene-d5	11.634	117	143401	25.000	ug/L	# 0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	81219	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.355	65	56563	19.158	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	76.640%
7) Chloroethane-d5	2.892	69	36504	18.357	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	73.440%
11) 1,1-Dichloroethene-d2	4.019	63	75990	17.588	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	70.360%
21) 2-Butanone-d5	7.080	46	19094	44.204	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	88.400%
24) Chloroform-d	7.653	84	110126	23.543	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	94.160%
26) 1,2-Dichloroethane-d4	8.305	65	54838	22.202	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	88.800%
32) Benzene-d6	8.275	84	224847	26.407	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	105.640%
36) 1,2-Dichloropropane-d6	9.275	67	59508	25.762	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	103.040%
41) Toluene-d8	10.323	98	223729	26.372	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	105.480%
43) trans-1,3-Dichloroprop...	10.579	79	23422	20.838	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	83.360%
47) 2-Hexanone-d5	10.927	63	16001	45.676	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	91.360%
56) 1,1,2,2-Tetrachloroeth...	12.695	84	44114	23.165	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	92.680%
66) 1,2-Dichlorobenzene-d4	13.853	152	82271	26.489	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	105.960%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW010722\
Data File : VW021770.D
Acq On : 07 Jan 2022 10:42
Operator : SY/VA
Sample : VW0107SBL01
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VBLK437

Quant Time: Jan 07 22:01:23 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM122821SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Jan 07 21:59:54 2022
Response via : Initial Calibration

