

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW010722\
 Data File : VW021782.D
 Acq On : 07 Jan 2022 17:54
 Operator : SY/VA
 Sample : N1074-05
 Misc : 5.81g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BGLK4

Quant Time: Jan 07 22:03:26 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM122821SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Jan 07 21:59:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	213676	25.000	ug/L	# 0.00
28) Chlorobenzene-d5	11.634	117	206808	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	116774	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	55449	13.161	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	52.640%
7) Chloroethane-d5	2.885	69	39801	14.027	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	56.120%
11) 1,1-Dichloroethene-d2	4.019	63	77214	12.524	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	50.080%
21) 2-Butanone-d5	7.080	46	20875	33.868	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	67.740%
24) Chloroform-d	7.653	84	121520	18.206	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	72.840%
26) 1,2-Dichloroethane-d4	8.305	65	60648	17.208	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	68.840%#
32) Benzene-d6	8.275	84	248468	20.234	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	80.920%
36) 1,2-Dichloropropane-d6	9.274	67	66584	19.988	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	79.960%
41) Toluene-d8	10.323	98	247910	20.262	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	81.040%
43) trans-1,3-Dichloroprop...	10.579	79	26976	16.642	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	66.560%
47) 2-Hexanone-d5	10.921	63	18727	37.068	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	74.140%
56) 1,1,2,2-Tetrachloroeth...	12.689	84	52405	19.082	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	76.320%
66) 1,2-Dichlorobenzene-d4	13.853	152	91008	20.380	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	81.520%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW010722\
Data File : VW021782.D
Acq On : 07 Jan 2022 17:54
Operator : SY/VA
Sample : N1074-05
Misc : 5.81g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BGLK4

Quant Time: Jan 07 22:03:26 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM122821SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Jan 07 21:59:54 2022
Response via : Initial Calibration

