

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW010722\
 Data File : VW021791.D
 Acq On : 07 Jan 2022 22:18
 Operator : SY/VA
 Sample : N1074-01
 Misc : 6.26g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BGLK0

Quant Time: Jan 07 23:04:48 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM122821SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Jan 07 22:46:04 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.848	114	185861	25.000	ug/L	# 0.00
28) Chlorobenzene-d5	11.634	117	188101	25.000	ug/L	# 0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	104797	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.355	65	48149	13.139	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	52.560%
7) Chloroethane-d5	2.885	69	35463	14.368	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	57.480%
11) 1,1-Dichloroethene-d2	4.019	63	64880	12.098	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	48.400%
21) 2-Butanone-d5	7.086	46	23111	43.107	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	86.220%
24) Chloroform-d	7.653	84	103899	17.896	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	71.600%
26) 1,2-Dichloroethane-d4	8.311	65	56150	18.316	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	73.280%
32) Benzene-d6	8.275	84	212097	18.990	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	75.960%
36) 1,2-Dichloropropane-d6	9.274	67	58195	19.207	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	76.840%
41) Toluene-d8	10.323	98	209269	18.805	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	75.240%
43) trans-1,3-Dichloroprop...	10.579	79	24181	16.401	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	65.600%
47) 2-Hexanone-d5	10.927	63	20988	45.675	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	91.340%
56) 1,1,2,2-Tetrachloroeth...	12.688	84	50390	20.173	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	80.680%
66) 1,2-Dichlorobenzene-d4	13.853	152	82526	20.593	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	82.360%
Target Compounds						
					Qvalue	
34) Trichloroethene	9.098	95	19828	6.086	ug/L	87
46) Tetrachloroethene	10.866	164	84123	27.572	ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW010722\
Data File : VW021791.D
Acq On : 07 Jan 2022 22:18
Operator : SY/VA
Sample : N1074-01
Misc : 6.26g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BGLK0

Quant Time: Jan 07 23:04:48 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM122821SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Jan 07 22:46:04 2022
Response via : Initial Calibration

