

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW010722\
 Data File : VW021780.D
 Acq On : 07 Jan 2022 17:06
 Operator : SY/VA
 Sample : N1053-22
 Misc : 7.19g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BGLJ9

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 01/10/2022
 Supervised By :Mahesh Dadoda 01/10/2022

Quant Time: Jan 07 22:03:05 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM122821SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Jan 07 21:59:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	38839	25.000	ug/L	# 0.00
28) Chlorobenzene-d5	11.628	117	39121	25.000	ug/L	# 0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	20280	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.355	65	10983	14.342	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	57.360%		
7) Chloroethane-d5	2.885	69	10779	20.899	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	83.600%		
11) 1,1-Dichloroethene-d2	4.025	63	15473	13.807	ug/L	0.00
Spiked Amount 25.000	Range 45 - 110		Recovery =	55.240%		
21) 2-Butanone-d5	7.080	46	4642m	41.434	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery =	82.860%		
24) Chloroform-d	7.653	84	22869	18.850	ug/L	0.00
Spiked Amount 25.000	Range 40 - 150		Recovery =	75.400%		
26) 1,2-Dichloroethane-d4	8.305	65	12636	19.724	ug/L	0.00
Spiked Amount 25.000	Range 70 - 130		Recovery =	78.880%		
32) Benzene-d6	8.275	84	47847	20.598	ug/L	0.00
Spiked Amount 25.000	Range 20 - 135		Recovery =	82.400%		
36) 1,2-Dichloropropane-d6	9.274	67	12801	20.314	ug/L	0.00
Spiked Amount 25.000	Range 70 - 120		Recovery =	81.240%		
41) Toluene-d8	10.323	98	45807	19.792	ug/L	0.00
Spiked Amount 25.000	Range 30 - 130		Recovery =	79.160%		
43) trans-1,3-Dichloroprop...	10.579	79	4637	15.122	ug/L	0.00
Spiked Amount 25.000	Range 30 - 135		Recovery =	60.480%		
47) 2-Hexanone-d5	10.927	63	3051	31.925	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery =	63.840%		
56) 1,1,2,2-Tetrachloroeth...	12.688	84	9992	19.233	ug/L	0.00
Spiked Amount 25.000	Range 45 - 120		Recovery =	76.920%		
66) 1,2-Dichlorobenzene-d4	13.847	152	17120	22.076	ug/L	0.00
Spiked Amount 25.000	Range 75 - 120		Recovery =	88.320%		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW010722\
Data File : VW021780.D
Acq On : 07 Jan 2022 17:06
Operator : SY/VA
Sample : N1053-22
Misc : 7.19g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BGLJ9

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 01/10/2022
Supervised By :Mahesh Dadoda 01/10/2022

Quant Time: Jan 07 22:03:05 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM122821SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Jan 07 21:59:54 2022
Response via : Initial Calibration

