

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW011420\
 Data File : VW014640.D
 Acq On : 14 Jan 2020 18:26
 Operator : SY/VA
 Sample : L1109-01
 Misc : 5.03G/10ML/MSVOA W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 GASG2

Quant Time: Jan 15 02:50:17 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM122719S.M
 Quant Title : VOC Analysis
 QLast Update : Wed Jan 15 01:48:42 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.84	114	1217780	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.63	117	1020528	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.55	152	374388	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.35	65	397360	23.15	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	92.60%
7) Chloroethane-d5	2.89	69	307170	24.29	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	97.16%
10) 1,1-Dichloroethene-d2	4.01	63	561366	17.88	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	71.52%
20) 2-Butanone-d5	7.07	46	169290	41.44	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	82.88%
24) Chloroform-d	7.65	84	680946	24.31	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	97.24%
26) 1,2-Dichloroethane-d4	8.30	65	342654	23.57	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	94.28%
29) Benzene-d6	8.27	84	1472097	27.13	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	108.52%
33) 1,2-Dichloropropane-d6	9.27	67	442725	27.13	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	108.52%
37) Toluene-d8	10.32	98	1309534	26.62	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	106.48%
38) trans-1,3-Dichloropropene-	10.58	79	160286	23.78	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	95.12%
39) 2-Hexanone-d5	10.92	63	131482	47.12	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	94.24%
48) 1,1,2,2-Tetrachloroethane-	12.69	84	282710	23.07	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	92.28%
61) 1,2-Dichlorobenzene-d4	13.85	152	321719	26.08	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	104.32%

Target Compounds

					Ovalue
13) Acetone	4.12	43	23848	4.217 ug/L	94
16) Methylene chloride	4.92	84	52574	2.648 ug/L	96
34) Benzene	8.32	78	43961	0.716 ug/L	100
44) Toluene	10.38	91	103071	1.606 ug/L	95
54) m,p-Xylene	11.83	106	16582	0.621 ug/L	88
55) o-xylene	12.16	106	7184	0.285 ug/L	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW011420\
Data File : VW014640.D
Acq On : 14 Jan 2020 18:26
Operator : SY/VA
Sample : L1109-01
Misc : 5.03G/10ML/MSVOA W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
GASG2

Quant Time: Jan 15 02:50:17 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM122719S.M
Quant Title : VOC Analysis
QLast Update : Wed Jan 15 01:48:42 2020
Response via : Initial Calibration

