

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW011622\
 Data File : VW021942.D
 Acq On : 15 Jan 2022 14:17
 Operator : SY/VA
 Sample : VW0116SBL01
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VBLK445

Quant Time: Jan 17 06:39:43 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM011122SMA.M
 Quant Title : SFAM01.0
 QLast Update : Mon Jan 17 06:38:36 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.848	114	140593	25.000	ug/L	# 0.00
28) Chlorobenzene-d5	11.628	117	138744	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	79145	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.355	65	47762	18.278	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	73.120%
7) Chloroethane-d5	2.891	69	35479	15.035	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	60.160%
11) 1,1-Dichloroethene-d2	4.025	63	67452	16.350	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	65.400%
21) 2-Butanone-d5	7.080	46	16303	44.201	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	88.400%
24) Chloroform-d	7.653	84	93253	20.531	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	82.120%
26) 1,2-Dichloroethane-d4	8.311	65	45925	20.938	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	83.760%
32) Benzene-d6	8.275	84	190479	21.158	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	84.640%
36) 1,2-Dichloropropane-d6	9.274	67	50261	21.181	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	84.720%
41) Toluene-d8	10.323	98	190499	20.991	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	83.960%
43) trans-1,3-Dichloroprop...	10.579	79	19852	19.705	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	78.800%
47) 2-Hexanone-d5	10.927	63	14177	44.031	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	88.060%
56) 1,1,2,2-Tetrachloroeth...	12.688	84	36026	19.416	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	77.680%
66) 1,2-Dichlorobenzene-d4	13.853	152	71724	22.040	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	88.160%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW011622\
Data File : VW021942.D
Acq On : 15 Jan 2022 14:17
Operator : SY/VA
Sample : VW0116SBL01
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VBLK445

Quant Time: Jan 17 06:39:43 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM011122SMA.M
Quant Title : SFAM01.0
QLast Update : Mon Jan 17 06:38:36 2022
Response via : Initial Calibration

