

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW011622\
 Data File : VW021947.D
 Acq On : 15 Jan 2022 16:16
 Operator : SY/VA
 Sample : N1168-05
 Misc : 6.85g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BGLX1

Quant Time: Jan 17 06:40:33 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM011122SMA.M
 Quant Title : SFAM01.0
 QLast Update : Mon Jan 17 06:38:36 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	184461	25.000	ug/L	# 0.00
28) Chlorobenzene-d5	11.628	117	180258	25.000	ug/L	# 0.00
58) 1,4-Dichlorobenzene-d4	13.560	152	103649	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.355	65	40940	11.941	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	47.760%
7) Chloroethane-d5	2.885	69	32883	10.621	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	42.480%
11) 1,1-Dichloroethene-d2	4.019	63	61285	11.322	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	45.280%
21) 2-Butanone-d5	7.086	46	16781	34.677	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	69.360%
24) Chloroform-d	7.653	84	95995	16.109	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	64.440%
26) 1,2-Dichloroethane-d4	8.311	65	46786	16.258	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	65.040%#
32) Benzene-d6	8.275	84	193307	16.527	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	66.120%
36) 1,2-Dichloropropane-d6	9.275	67	51889	16.831	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	67.320%#
41) Toluene-d8	10.323	98	184301	15.631	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	62.520%
43) trans-1,3-Dichloroprop...	10.579	79	19405	14.825	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	59.320%
47) 2-Hexanone-d5	10.927	63	14295	34.172	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	68.340%
56) 1,1,2,2-Tetrachloroeth...	12.689	84	38599	16.012	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	64.040%
66) 1,2-Dichlorobenzene-d4	13.853	152	73758	17.306	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	69.240%#

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW011622\
Data File : VW021947.D
Acq On : 15 Jan 2022 16:16
Operator : SY/VA
Sample : N1168-05
Misc : 6.85g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BGLX1

Quant Time: Jan 17 06:40:33 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM011122SMA.M
Quant Title : SFAM01.0
QLast Update : Mon Jan 17 06:38:36 2022
Response via : Initial Calibration

