

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW011722\
 Data File : VW021968.D
 Acq On : 17 Jan 2022 12:34
 Operator : SY/VA
 Sample : N1129-07
 Misc : 5.66g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 DBM00

Quant Time: Jan 18 00:37:42 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM011122SMA.M
 Quant Title : SFAM01.0
 QLast Update : Tue Jan 18 00:36:09 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	153274	25.000	ug/L	# 0.00
28) Chlorobenzene-d5	11.628	117	149122	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	87525	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.355	65	54841	19.250	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	77.000%
7) Chloroethane-d5	2.885	69	36919	14.351	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	57.400%
11) 1,1-Dichloroethene-d2	4.013	63	73150	16.264	ug/L	-0.01
Spiked Amount	25.000	Range	45 - 110	Recovery	=	65.040%
21) 2-Butanone-d5	7.086	46	19661	48.896	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	97.800%
24) Chloroform-d	7.647	84	105039	21.213	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	84.840%
26) 1,2-Dichloroethane-d4	8.305	65	53995	22.581	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	90.320%
32) Benzene-d6	8.275	84	216743	22.400	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	89.600%
36) 1,2-Dichloropropane-d6	9.274	67	58854	23.077	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	92.320%
41) Toluene-d8	10.323	98	218050	22.354	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	89.400%
43) trans-1,3-Dichloroprop...	10.579	79	23525	21.725	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	86.920%
47) 2-Hexanone-d5	10.921	63	17428	50.360	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	100.720%
56) 1,1,2,2-Tetrachloroeth...	12.688	84	44752	22.440	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	89.760%
66) 1,2-Dichlorobenzene-d4	13.853	152	83439	23.185	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	92.720%
Target Compounds						
53) m,p-Xylene	11.835	106	2999	0.630	ug/L	Qvalue 80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW011722\
Data File : VW021968.D
Acq On : 17 Jan 2022 12:34
Operator : SY/VA
Sample : N1129-07
Misc : 5.66g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
DBM00

Quant Time: Jan 18 00:37:42 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM011122SMA.M
Quant Title : SFAM01.0
QLast Update : Tue Jan 18 00:36:09 2022
Response via : Initial Calibration

