

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW011722\
 Data File : VW021973.D
 Acq On : 17 Jan 2022 14:34
 Operator : SY/VA
 Sample : N1129-02
 Misc : 6.78g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 DBLZ7

Quant Time: Jan 18 00:39:14 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM011122SMA.M
 Quant Title : SFAM01.0
 QLast Update : Tue Jan 18 00:36:09 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	147233	25.000	ug/L	# 0.00
28) Chlorobenzene-d5	11.628	117	130281	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	52786	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	42146	15.401	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	61.600%
7) Chloroethane-d5	2.885	69	33349	13.495	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	54.000%
11) 1,1-Dichloroethene-d2	4.013	63	56747	13.135	ug/L	-0.01
Spiked Amount	25.000	Range	45 - 110	Recovery	=	52.520%
21) 2-Butanone-d5	7.086	46	15419	39.919	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	79.840%
24) Chloroform-d	7.647	84	93039	19.560	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	78.240%
26) 1,2-Dichloroethane-d4	8.305	65	48439	21.089	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	84.360%
32) Benzene-d6	8.275	84	180298	21.328	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	85.320%
36) 1,2-Dichloropropane-d6	9.274	67	51249	23.001	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	92.000%
41) Toluene-d8	10.323	98	164899	19.350	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	77.400%
43) trans-1,3-Dichloroprop...	10.579	79	16392	17.327	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	69.320%
47) 2-Hexanone-d5	10.927	63	12617	41.731	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	83.460%
56) 1,1,2,2-Tetrachloroeth...	12.688	84	35413	20.325	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	81.320%
66) 1,2-Dichlorobenzene-d4	13.847	152	42855	19.744	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	78.960%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW011722\
Data File : VW021973.D
Acq On : 17 Jan 2022 14:34
Operator : SY/VA
Sample : N1129-02
Misc : 6.78g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
DBLZ7

Quant Time: Jan 18 00:39:14 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM011122SMA.M
Quant Title : SFAM01.0
QLast Update : Tue Jan 18 00:36:09 2022
Response via : Initial Calibration

